



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106264

Slope : 98.5

pH Meter ID : JIGNESH

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	11/13/2019	16:10
2	CAL2	1	Water	NA	NA	20.4	7.01	11/13/2019	16:11
3	CAL3	1	Water	NA	NA	20.4	10.03	11/13/2019	16:12
4	ICV	1	Water	NA	NA	20.3	7.01	11/13/2019	16:15
5	CCV1	1	Water	NA	NA	20.1	2.02	11/13/2019	16:16
6	K5814-02	1	Water	NA	NA	11.5	6.70	11/13/2019	16:17
7	K5814-04	1	Water	NA	NA	12.6	6.34	11/13/2019	16:19
8	K5814-06	1	Water	NA	NA	11.9	6.90	11/13/2019	16:20
9	K5814-08	1	Water	NA	NA	13.7	6.31	11/13/2019	16:21
10	K5814-10	1	Water	NA	NA	13.6	6.90	11/13/2019	16:22
11	K5822-01	1	Water	NA	NA	14.8	6.95	11/13/2019	16:25
12	K5822-01DUP	1	Water	NA	NA	14.9	6.96	11/13/2019	16:26
13	CCV2	1	Water	NA	NA	20.1	12.02	11/13/2019	16:30

WORKLIST(Hardcopy Internal Chain)

LB 106264

WorkList Name : PH W K5814

WorkList ID : 133220

Department : Wet-Chemistry

Date : 11-13-2019 15:18:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/13/2019	Water	K5814-02 -C	pH	Cool 4 deg C	PSEG01	N62	30557	11/13/2019	9040C
11/13/2019	Water	K5814-04 -C	pH	Cool 4 deg C	PSEG01	N62	40189	11/13/2019	9040C
11/13/2019	Water	K5814-06 -C	pH	Cool 4 deg C	PSEG01	N62	30663	11/13/2019	9040C
11/13/2019	Water	K5814-08 -C	pH	Cool 4 deg C	PSEG01	N62	30662	11/13/2019	9040C
11/13/2019	Water	K5814-10 -C	pH	Cool 4 deg C	PSEG01	N62	30535	11/13/2019	9040C
11/13/2019	Water	K5822-01 -P	pH	Cool 4 deg C	PSEG01	N62	N35868	11/13/2019	9040C

Date/Time 11-13-19 15:20
 Received by: Jn
 Relinquished by: Jn

Date/Time 11-13-19 17:00
 Received by: Jn
 Relinquished by: Jn