



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106265

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	11/13/2019	16:10
2	CAL2	1	Water	NA	NA	20.4	7.01	11/13/2019	16:11
3	CAL3	1	Water	NA	NA	20.4	10.03	11/13/2019	16:12
4	ICV	1	Water	NA	NA	20.3	7.01	11/13/2019	16:15
5	CCV1	1	Water	NA	NA	20.1	2.02	11/13/2019	16:16
6	K5822-02	1	Water	NA	NA	14.8	6.95	11/13/2019	16:25
7	K5822-02DUP	1	Water	NA	NA	14.9	6.96	11/13/2019	16:26
8	CCV2	1	Water	NA	NA	20.1	12.02	11/13/2019	16:30

2B106265

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K5822 WorkList ID : 133229 Department : Wet-Chemistry Date : 11-13-2019 15:19:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/13/2019	Water	K5822-02	-1	Corrosivity	PSEG01	N62	N35868	11/13/2019	9040C

Date/Time 11-13-19 15:19:02
Received by: Jo
Relinquished by: Jo

Date/Time 11-13-19 17:10:00
Received by: Jo
Relinquished by: Jo