



PERCENT SOLID

Supervisor: apatel
Analyst: Dodley
Date: 11/15/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 18:18
In Date: 11/14/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 11/15/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106287

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
K5665-01	HR-05-111419	33	1.14	9.87	9.16	91.9	
K5665-02	HR-05-111419-1	34	1.14	9.62	8.81	90.4	
K5665-03	HR-05-111419-2	35	1.13	9.59	8.95	92.4	
K5671-06	OR-03-111419	36	1.14	9.6	9.01	93.0	
K5671-07	OR-03-111419	37	1.14	9.93	9.36	93.5	
K5671-08	OR-03-111419-1	38	1.14	9.97	9.32	92.6	
K5671-09	OR-03-111419-2	39	1.14	9.8	9.16	92.6	
K5832-01	10-(2-4)	1	1.15	9.69	9.07	92.7	
K5832-02	10-(4-6)	2	1.14	9.62	9.01	92.8	
K5832-03	10-(6-8)	3	1.13	9.71	9.13	93.2	
K5832-04	9-(2-4)	4	1.15	9.72	8.7	88.1	
K5832-05	9-(4-6)	5	1.14	9.96	8.86	87.5	
K5832-06	9-(6-8)	6	1.16	9.81	8.84	88.8	
K5832-07	9-(8-10)	7	1.16	9.8	8.93	89.9	
K5832-08	8-(4-6)	8	1.13	9.9	9.00	89.7	
K5832-09	8-(6-8)	9	1.12	9.65	8.44	85.8	
K5832-10	8-(8-10)	10	1.15	9.79	8.97	90.5	
K5832-11	6-(6-8)	11	1.15	9.65	9.19	94.6	
K5832-12	6-(8-10)	12	1.15	9.82	9.34	94.5	
K5832-13	6-(10-12)	13	1.14	9.84	9.49	96.0	
K5832-14	6-(12-15)	14	1.15	9.71	8.93	90.9	
K5832-15	5-(2-4)	15	1.14	9.93	9.06	90.1	
K5832-16	5-(4-6)	16	1.15	9.57	8.72	89.9	
K5832-17	5-(6-8)	17	1.15	9.69	8.8	89.6	
K5832-18	5-(8-10)	18	1.14	9.8	8.73	87.6	
K5832-19	5-(10-12)	19	1.14	9.93	9.29	92.7	
K5832-20	5-(12-14)	20	1.15	9.8	9.42	95.6	
K5833-01	8-(10-12)	21	1.14	9.97	9.52	94.9	



PERCENT SOLID

Supervisor: apatel
Analyst: Dodley
Date: 11/15/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 18:18
In Date: 11/14/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 11/15/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106287

Table with 8 columns: Lab ID, Client SampleID, Dish #, Dish Wt (g) (A), Dish + Sample Wt (g) (B), Dish+Dry Sample Wt (g) (C), % Solid, Comments. It contains 35 rows of data for various samples.



PERCENT SOLID

Supervisor: apatel
Analyst: Dodley
Date: 11/15/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 18:18
In Date: 11/14/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 11/15/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106287

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
K5863-05	TP-1-COMP	46	1.14	9.7	8.9	90.7	
K5864-01	SB-01-7.5-8.0	59	1.14	9.93	9.34	93.3	
K5864-02	SB-01-COMP	60	1.14	9.93	9.34	93.3	
K5864-03	SB-02-7.5-8.0	61	1.14	9.79	9.14	92.5	
K5864-04	SB-02-COMP	62	1.14	9.79	9.14	92.5	
K5864-05	SB-03-7.5-8.0	63	1.15	9.81	9.04	91.1	
K5864-06	SB-03-COMP	64	1.15	9.81	9.04	91.1	
K5864-07	SB-04-7.5-8.0	65	1.15	9.73	8.59	86.7	
K5864-08	SB-04-COMP	66	1.15	9.73	8.59	86.7	
K5864-09	SB-05-7.5-8.0	67	1.15	9.58	8.79	90.6	
K5864-10	SB-05-COMP	68	1.15	9.58	8.79	90.6	
K5866-01	OIL-TEST-UNDER-BRIDGE	52	1.13	9.73	9.37	95.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$