



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106291

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	11/14/2019	16:30
2	CAL2	1	Water	NA	NA	20.4	7.01	11/14/2019	16:31
3	CAL3	1	Water	NA	NA	20.4	10.03	11/14/2019	16:33
4	ICV	1	Water	NA	NA	20.4	7.00	11/14/2019	16:35
5	CCV1	1	Water	NA	NA	20.1	2.02	11/14/2019	16:36
6	K5852-04	1	Water	NA	NA	19.2	9.23	11/14/2019	16:37
7	K5852-04DUP	1	Water	NA	NA	19.1	9.24	11/14/2019	16:38
8	K5857-05	1	Water	NA	NA	20.1	5.14	11/14/2019	16:40
9	K5857-06	1	Water	NA	NA	16.2	4.78	11/14/2019	16:41
10	K5857-07	1	Water	NA	NA	16.3	5.85	11/14/2019	16:42
11	K5857-08	1	Water	NA	NA	16.5	4.95	11/14/2019	16:44
12	K5860-07	1	Water	NA	NA	15.9	5.05	11/14/2019	16:45
13	K5860-08	1	Water	NA	NA	17.0	5.17	11/14/2019	16:47
14	CCV2	1	Water	NA	NA	20.1	12.01	11/14/2019	16:48

✓ 106291

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH W K5852 WorkList ID : 133292 Department : Wet-Chemistry Date : 11-14-2019 16:14:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/14/2019	Water	K5852-04	pH	Cool 4 deg C	PSEG01	N63	20924	11/14/2019	9040C
11/14/2019	Water	K5857-05	pH	Cool 4 deg C	PSEG01	N63	20936	11/14/2019	9040C
11/14/2019	Water	K5857-06	pH	Cool 4 deg C	PSEG01	N63	20964	11/14/2019	9040C
11/14/2019	Water	K5857-07	pH	Cool 4 deg C	PSEG01	N63	20965	11/14/2019	9040C
11/14/2019	Water	K5857-08	pH	Cool 4 deg C	PSEG01	N63	20981	11/14/2019	9040C
11/14/2019	Water	K5860-07	pH	Cool 4 deg C	PSEG01	N62	20983	11/14/2019	9040C
11/14/2019	Water	K5860-08	pH	Cool 4 deg C	PSEG01	N62	32007	11/14/2019	9040C

Date/Time 11-14-19 16:20
Received by: fo
Relinquished by: cd

Date/Time 11-14-19 18:30
Received by: cd
Relinquished by: fo