



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106292

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	11/14/2019	16:30
2	CAL2	1	Water	NA	NA	20.4	7.01	11/14/2019	16:31
3	CAL3	1	Water	NA	NA	20.4	10.03	11/14/2019	16:33
4	ICV	1	Water	NA	NA	20.4	7.00	11/14/2019	16:35
5	CCV1	1	Water	NA	NA	20.3	2.01	11/14/2019	15:00
6	K5852-01	1	Solid	20.02	20	19.8	9.67	11/14/2019	17:01
7	K5852-03	1	Solid	20.03	20	20.2	5.54	11/14/2019	17:02
8	K5857-01	1	Solid	20.04	20	19.6	7.70	11/14/2019	17:05
9	K5860-01	1	Solid	20.02	20	19.8	6.14	11/14/2019	17:06
10	K5860-03	1	Solid	20.02	20	20.4	6.47	11/14/2019	17:07
11	K5860-09	1	Solid	20.03	20	19.1	6.54	11/14/2019	17:10
12	K5860-09DUP	1	Solid	20.04	20	19.0	6.55	11/14/2019	17:11
13	CCV2	1	Water	NA	NA	20.1	12.02	11/14/2019	17:12

2B 106292

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S K5852 WorkList ID : 133293 Department : Wet-Chemistry Date : 11-14-2019 16:15:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/14/2019	Solid	K5852-01 ✓	pH	Cool 4 deg C	PSEG01	N63	21970	11/14/2019	9045D
11/14/2019	Solid	K5852-03 ✓	pH	Cool 4 deg C	PSEG01	N63	20987	11/14/2019	9045D
11/14/2019	Solid	K5857-01 ✓	pH	Cool 4 deg C	PSEG01	N63	10327	11/14/2019	9045D
11/14/2019	Solid	K5860-01 ✓	pH	Cool 4 deg C	PSEG01	N62	20949-20950	11/14/2019	9045D
11/14/2019	Solid	K5860-03 ✓	pH	Cool 4 deg C	PSEG01	N62	20957-20958-20959-20960	11/14/2019	9045D
11/14/2019	Solid	K5860-09 ✓	pH	Cool 4 deg C	PSEG01	N62	20777	11/14/2019	9045D

Date/Time 11-14-19 16:20
 Received by: 77
 Relinquished by: 77

Date/Time 11-14-19 18:30
 Received by: 77
 Relinquished by: 77