

LB106298


7471B

INSTRUMENT ID CV1

Linear

 μ Abs.:

59783



A= 0.0000e+000

B= 1.6739e-004

C= -4.4130e-002

Rho= 0.9998909

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.033	-0.033	68	0.000	68				
0.2	0.200	0.187	-0.013	1383	0.0 %	1383				
2.5	2.500	2.585	0.085	15706	0.0 %	15706				
5.0	5.000	4.935	-0.065	29748	0.0 %	29748				
7.5	7.500	7.562	0.062	45439	0.0 %	45439				
10.0	10.000	9.963	-0.037	59783	0.0 %	59783				

LB106298 INSTRUMENT ID CV1

Method: 7471B		Operator: Admin		Date of Analysis: 15 Nov 2019 07:53:04					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	68	-	0.0000	7471B	PPB	15 Nov 2019 08:01:37	S	Std
0.2 - 1	50.2	1383	-	0.2000	7471B	PPB	15 Nov 2019 08:03:55	S	Std
2.5 - 1	52.5	15706	-	2.5000	7471B	PPB	15 Nov 2019 08:06:14	S	Std
5.0 - 1	55.0	29748	-	5.0000	7471B	PPB	15 Nov 2019 08:08:31	S	Std
7.5 - 1	57.5	45439	-	7.5000	7471B	PPB	15 Nov 2019 08:10:49	S	Std
10.0 - 1	510.0	59783	-	10.0000	7471B	PPB	15 Nov 2019 08:13:06	S	Std
ICV06 - 1	ICV06	22763	3.7662	-	7471B	PPB	15 Nov 2019 08:28:19	U	SMPL
ICB06 - 1	ICB06	0	-0.0441	-	7471B	PPB	15 Nov 2019 08:30:35	U	SMPL
CCV53 - 1	CCV53	29475	4.8897	-	7471B	PPB	15 Nov 2019 08:32:51	U	SMPL
CCB53 - 1	CCB53	-7	-0.0453	-	7471B	PPB	15 Nov 2019 08:35:07	U	SMPL
CRA - 1	CRA	1399	0.1901	-	7471B	PPB	15 Nov 2019 08:37:22	U	SMPL
HighStd - 1	HighStd	60488	10.0811	-	7471B	PPB	15 Nov 2019 08:39:39	U	SMPL
ChkStd - 1	ChkStd	41629	6.9242	-	7471B	PPB	15 Nov 2019 08:41:54	U	SMPL
PB124797BL - 1	PBS	-35	-0.0500	-	7471B	PPB	15 Nov 2019 08:44:12	U	SMPL
PB124797BS - 1	LCSS	20794	3.4366	-	7471B	PPB	15 Nov 2019 08:46:30	U	SMPL
K5668-05 - 1	HD-02-111319	845	0.0973	-	7471B	PPB	15 Nov 2019 08:48:47	U	SMPL
K5668-05DUP - 1	HD-02-111319DUP	1074	0.1356	-	7471B	PPB	15 Nov 2019 08:51:04	U	SMPL
K5668-05MS - 1	HD-02-111319MS	21958	3.6315	-	7471B	PPB	15 Nov 2019 08:53:22	U	SMPL
K5668-05MSD - 1	HD-02-111319MSD	21923	3.6256	-	7471B	PPB	15 Nov 2019 08:55:40	U	SMPL
K5806-01 - 1	STOCK-PILE	9441	1.5362	-	7471B	PPB	15 Nov 2019 08:57:56	U	SMPL
CCV54 - 1	CCV54	28126	4.6639	-	7471B	PPB	15 Nov 2019 09:00:11	U	SMPL
CCB54 - 1	CCB54	-55	-0.0533	-	7471B	PPB	15 Nov 2019 09:02:27	U	SMPL
K5807-01 - 1	P001-CC01-191111-01	400	0.0228	-	7471B	PPB	15 Nov 2019 09:04:43	U	SMPL
K5807-03 - 1	P001-CC02-191111-01	3286	0.5059	-	7471B	PPB	15 Nov 2019 09:06:59	U	SMPL
K5807-05 - 1	P001-CC03-191111-01	766	0.0841	-	7471B	PPB	15 Nov 2019 09:09:15	U	SMPL
K5807-07 - 1	P001-CC04-191111-01	325	0.0103	-	7471B	PPB	15 Nov 2019 09:11:31	U	SMPL
K5807-09 - 1	P001-CC05-191111-01	648	0.0643	-	7471B	PPB	15 Nov 2019 09:13:48	U	SMPL
K5807-11 - 1	P001-CC06-191111-01	989	0.1214	-	7471B	PPB	15 Nov 2019 09:16:05	U	SMPL
K5807-13 - 1	P001-CC07-191111-01	2272	0.3362	-	7471B	PPB	15 Nov 2019 09:18:22	U	SMPL
K5807-15 - 1	P001-CC08-191112-01	175	-0.0148	-	7471B	PPB	15 Nov 2019 09:20:39	U	SMPL
K5807-17 - 1	P001-CC09-191112-01	387	0.0207	-	7471B	PPB	15 Nov 2019 09:22:57	U	SMPL
K5807-19 - 1	P001-CC10-191112-01	449	0.0310	-	7471B	PPB	15 Nov 2019 09:25:13	U	SMPL
CCV55 - 1	CCV55	28088	4.6576	-	7471B	PPB	15 Nov 2019 09:27:29	U	SMPL
CCB55 - 1	CCB55	-22	-0.0478	-	7471B	PPB	15 Nov 2019 09:29:45	U	SMPL
K5807-21 - 1	P001-CC11-191112-01	394	0.0218	-	7471B	PPB	15 Nov 2019 09:32:01	U	SMPL
K5807-23 - 1	P001-CC12-191112-01	1175	0.1526	-	7471B	PPB	15 Nov 2019 09:34:17	U	SMPL
K5807-26 - 1	P001-CC13-191112-01	522	0.0432	-	7471B	PPB	15 Nov 2019 09:36:33	U	SMPL
K5813-01 - 1	SP-B	43047	7.1616	-	7471B	PPB	15 Nov 2019 09:38:49	U	SMPL
K5818-01 - 1	30525	-39	-0.0507	-	7471B	PPB	15 Nov 2019 09:41:06	U	SMPL
K5818-03 - 1	40180	144	-0.0200	-	7471B	PPB	15 Nov 2019 09:43:24	U	SMPL
K5826-03 - 1	TP-15-S	972	0.1186	-	7471B	PPB	15 Nov 2019 09:45:42	U	SMPL
K5826-07 - 1	TP-14B-S	374	0.0185	-	7471B	PPB	15 Nov 2019 09:47:59	U	SMPL
PB124822BL - 1	PBS	42	-0.0371	-	7471B	PPB	15 Nov 2019 09:50:17	U	SMPL
PB124822BS - 1	LCSS	21003	3.4716	-	7471B	PPB	15 Nov 2019 09:52:33	U	SMPL
CCV56 - 1	CCV56	30155	5.0036	-	7471B	PPB	15 Nov 2019 09:54:50	U	SMPL
CCB56 - 1	CCB56	-19	-0.0473	-	7471B	PPB	15 Nov 2019 09:57:06	U	SMPL
K5832-01 - 1	10-(2-4)	2341	0.3477	-	7471B	PPB	15 Nov 2019 09:59:23	U	SMPL
K5832-02 - 1	10-(4-6)	2555	0.3836	-	7471B	PPB	15 Nov 2019 10:01:40	U	SMPL
K5832-03 - 1	10-(6-8)	1826	0.2615	-	7471B	PPB	15 Nov 2019 10:03:56	U	SMPL
K5832-03DUP - 1	10-(6-8)DUP	1830	0.2622	-	7471B	PPB	15 Nov 2019 10:06:12	U	SMPL
K5832-03MS - 1	10-(6-8)MS	26581	4.4053	-	7471B	PPB	15 Nov 2019 10:08:28	U	SMPL
K5832-03MSD - 1	10-(6-8)MSD	26341	4.3651	-	7471B	PPB	15 Nov 2019 10:10:45	U	SMPL
K5832-04 - 1	9-(2-4)	2207	0.3253	-	7471B	PPB	15 Nov 2019 10:13:03	U	SMPL
K5832-05 - 1	9-(4-6)	4058	0.6351	-	7471B	PPB	15 Nov 2019 10:15:20	U	SMPL
K5832-06 - 1	9-(6-8)	12232	2.0034	-	7471B	PPB	15 Nov 2019 10:17:38	U	SMPL
K5832-07 - 1	9-(8-10)	7971	1.2901	-	7471B	PPB	15 Nov 2019 10:19:53	U	SMPL
CCV57 - 1	CCV57	30121	4.9979	-	7471B	PPB	15 Nov 2019 10:22:10	U	SMPL
CCB57 - 1	CCB57	-22	-0.0478	-	7471B	PPB	15 Nov 2019 10:24:27	U	SMPL
K5832-08 - 1	8-(4-6)	709	0.0746	-	7471B	PPB	15 Nov 2019 10:26:43	U	SMPL
K5832-09 - 1	8-(6-8)	1243	0.1639	-	7471B	PPB	15 Nov 2019 10:29:01	U	SMPL
K5832-10 - 1	8-(8-10)	2907	0.4425	-	7471B	PPB	15 Nov 2019 10:31:17	U	SMPL
K5832-11 - 1	6-(6-8)	604	0.0570	-	7471B	PPB	15 Nov 2019 10:33:34	U	SMPL
K5832-12 - 1	6-(8-10)	606	0.0573	-	7471B	PPB	15 Nov 2019 10:35:50	U	SMPL
K5832-13 - 1	6-(10-12)	480	0.0362	-	7471B	PPB	15 Nov 2019 10:38:08	U	SMPL
K5832-14 - 1	6-(12-15)	244	-0.0033	-	7471B	PPB	15 Nov 2019 10:40:25	U	SMPL
K5832-15 - 1	5-(2-4)	6315	1.0129	-	7471B	PPB	15 Nov 2019 10:42:42	U	SMPL
K5832-16 - 1	5-(4-6)	6624	1.0647	-	7471B	PPB	15 Nov 2019 10:45:01	U	SMPL
K5832-17 - 1	5-(6-8)	3673	0.5707	-	7471B	PPB	15 Nov 2019 10:47:17	U	SMPL
CCV58 - 1	CCV58	30286	5.0255	-	7471B	PPB	15 Nov 2019 10:49:34	U	SMPL
CCB58 - 1	CCB58	-21	-0.0476	-	7471B	PPB	15 Nov 2019 10:51:50	U	SMPL
K5832-18 - 1	5-(8-10)	3518	0.5448	-	7471B	PPB	15 Nov 2019 10:54:08	U	SMPL
K5832-19 - 1	5-(10-12)	1682	0.2374	-	7471B	PPB	15 Nov 2019 10:56:24	U	SMPL
K5832-20 - 1	5-(12-14)	339	0.0126	-	7471B	PPB	15 Nov 2019 10:58:41	U	SMPL
PB124823BL - 1	PBS	15	-0.0416	-	7471B	PPB	15 Nov 2019 11:00:57	U	SMPL
PB124823BS - 1	LCSS	21981	3.6353	-	7471B	PPB	15 Nov 2019 11:03:14	U	SMPL
K5833-01 - 1	8-(10-12)	249	-0.0024	-	7471B	PPB	15 Nov 2019 11:05:30	U	SMPL
K5833-01DUP - 1	8-(10-12)DUP	270	0.0011	-	7471B	PPB	15 Nov 2019 11:07:47	U	SMPL
K5833-01MS - 1	8-(10-12)MS	21555	3.5640	-	7471B	PPB	15 Nov 2019 11:10:04	U	SMPL
K5833-01MSD - 1	8-(10-12)MSD	21427	3.5426	-	7471B	PPB	15 Nov 2019 11:12:22	U	SMPL
K5833-02 - 1	8-(12-13)	1305	0.1743	-	7471B	PPB	15 Nov 2019 11:14:39	U	SMPL
CCV59 - 1	CCV59	29243	4.8509	-	7471B	PPB	15 Nov 2019 11:16:55	U	SMPL

LB106298 INSTRUMENT ID CV1

Method: 7471B		Operator: Admin		Date of Analysis: 15 Nov 2019 07:53:04					
Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
CCB59 - 1	CCB59	-15	-0.0466		-7471B	PPB	15 Nov 2019 11:19:12	U	SMPL
K5833-03 - 1	7-(2-4)	11598	1.8973		-7471B	PPB	15 Nov 2019 11:21:29	U	SMPL
K5833-04 - 1	7-(4-6)	7838	1.2679		-7471B	PPB	15 Nov 2019 11:23:46	U	SMPL
K5833-05 - 1	7-(6-8)	10245	1.6708		-7471B	PPB	15 Nov 2019 11:26:02	U	SMPL
K5833-06 - 1	7-(8-10)	260	-0.0006		-7471B	PPB	15 Nov 2019 11:28:19	U	SMPL
K5833-07 - 1	7-(10-12)	9334	1.5183		-7471B	PPB	15 Nov 2019 11:30:36	U	SMPL
K5833-08 - 1	7-(12-14)	971	0.1184		-7471B	PPB	15 Nov 2019 11:32:53	U	SMPL
K5833-09 - 1	6-(2-4)	5216	0.8290		-7471B	PPB	15 Nov 2019 11:35:10	U	SMPL
K5833-10 - 1	6-(4-6)	6237	0.9999		-7471B	PPB	15 Nov 2019 11:37:27	U	SMPL
K5833-12 - 1	8-(2-4)	591	0.0548		-7471B	PPB	15 Nov 2019 11:39:44	U	SMPL
K5668-05LX5 - 1		163	-0.0168		-7471B	PPB	15 Nov 2019 11:42:01	U	SMPL
CCV60 - 1	CCV60	29672	4.9227		-7471B	PPB	15 Nov 2019 11:50:30	U	SMPL
CCB60 - 1	CCB60	3	-0.0436		-7471B	PPB	15 Nov 2019 11:52:47	U	SMPL
K5832-03LX5 - 1		170	-0.0157		-7471B	PPB	15 Nov 2019 12:06:06	U	SMPL
K5833-01LX5 - 1		68	-0.0327		-7471B	PPB	15 Nov 2019 12:08:24	U	SMPL
CCV61 - 1	CCV61	29123	4.8308		-7471B	PPB	15 Nov 2019 12:10:40	U	SMPL
CCB61 - 1	CCB61	18	-0.0411		-7471B	PPB	15 Nov 2019 12:12:57	U	SMPL