

LB106309


ISM02.4

INSTRUMENT ID CV1

Linear

 μ Abs.:

47364



A= 0.0000e+000

B= 2.1234e-004 slope

C= 4.2353e-002 y-intercept

Rho= 0.9996988

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	D%).
0.00	0.000	0.042	0.042	0	0.000	0					
0.05	0.050										
0.20	0.200	0.235	0.035	909	0.0 %	909					17
2.50	2.500	2.537	0.037	11747	0.0 %	11747					1
5.00	5.000	4.815	-0.185	22479	0.0 %	22479					-3
7.50	7.500	7.471	-0.029	34983	0.0 %	34983					0
10.0	10.000	10.100	0.100	47364	0.0 %	47364					0

9
10/15/19

INSTRUMENT ID CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0.0		0 -		0 ISM02.4	PPB	11/15/2019 13:20	Std
0.2 S0.2		909 -		0.2 ISM02.4	PPB	11/15/2019 13:22	Std
2.5 S2.5		11747 -		2.5 ISM02.4	PPB	11/15/2019 13:25	Std
5 S5.0		22479 -		5 ISM02.4	PPB	11/15/2019 13:27	Std
7.5 S7.5		34983 -		7.5 ISM02.4	PPB	11/15/2019 13:29	Std
10 S10.0		47364 -		10 ISM02.4	PPB	11/15/2019 13:32	Std
ICV	ICV	18329	3.9343 -	ISM02.4	PPB	11/15/2019 13:34	SMPL
ICB	ICB	-49	0.0319 -	ISM02.4	PPB	11/15/2019 13:37	SMPL
CCV062	CCV062	23970	5.1321 -	ISM02.4	PPB	11/15/2019 13:39	SMPL
CCB062	CCB062	-144	0.0118 -	ISM02.4	PPB	11/15/2019 13:41	SMPL
PB124824E	PBW824	12	0.0449 -	ISM02.4	PPB	11/15/2019 13:43	SMPL
K5816-01	MA5476	13	0.0451 -	ISM02.4	PPB	11/15/2019 13:46	SMPL
K5816-02	MA5476D	8	0.0441 -	ISM02.4	PPB	11/15/2019 13:48	SMPL
K5816-03	MA5476S	3121	0.7051 -	ISM02.4	PPB	11/15/2019 13:50	SMPL
K5817-01	MA5477	7	0.0438 -	ISM02.4	PPB	11/15/2019 13:53	SMPL
K5817-02	MA5477D	0	0.0424 -	ISM02.4	PPB	11/15/2019 13:55	SMPL
K5817-03	MA5477S	4403	0.9773 -	ISM02.4	PPB	11/15/2019 13:57	SMPL
CCV063	CCV063	24020	5.1427 -	ISM02.4	PPB	11/15/2019 13:59	SMPL
CCB063	CCB063	-25	0.037 -	ISM02.4	PPB	11/15/2019 14:02	SMPL