

CB106317

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AC

Instrument ID : Konelab

11/15/2019 14:05

Test: Phenolics-

Sample Id	Result <i>mg/kg</i>	Dil. 1 +	Response	Errors
ICV1	1.005	0.0	0.109	
ICB1	-0.002	0.0	0.006	
CCV1	0.964	0.0	0.104	
CCB1	0.007	0.0	0.007	
PB124760BL	0.005	0.0	0.007	
PB124760BS	0.992	0.0	0.107	
K5804-01	0.011	0.0	0.007	
K5804-04	0.005	0.0	0.006	
K5804-04DUP	0.003	0.0	0.006	
K5804-04MS	0.863	0.0	0.094	
K5804-04MSD	0.859	0.0	0.094	
K5828-01	0.005	0.0	0.007	
K5828-02	0.005	0.0	0.006	
K5828-03	0.013	0.0	0.007	
CCV2	0.959	0.0	0.104	
CCB2	0.012	0.0	0.007	

N	16
Mean	0.357
SD	0.4684
CV%	131.33

Aquakem v. 7.2AQ1

AC

Results from time period:

Fri Nov 15 13:28:16 2019

Fri Nov 15 14:01:57 2019

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0518	mg/l	11/15/2019 13:28:16	
0.1PPM	A	Phenolics-	P	0.1029	mg/l	11/15/2019 13:28:17	
0.25PPM	A	Phenolics-	P	0.2512	mg/l	11/15/2019 13:28:18	
0.50PPM	A	Phenolics-	P	0.5011	mg/l	11/15/2019 13:28:19	
1.0PPM	A	Phenolics-	P	0.9872	mg/l	11/15/2019 13:28:20	
2.0PPM	A	Phenolics-	P	2.0058	mg/l	11/15/2019 13:28:21	
ICV1	S	Phenolics-	P	1.0054	mg/l	11/15/2019 13:56:31	
ICB1	S	Phenolics-	P	-0.0025	mg/l	11/15/2019 13:56:32	
CCV1	S	Phenolics-	P	0.9638	mg/l	11/15/2019 13:56:33	
CCB1	S	Phenolics-	P	0.0074	mg/l	11/15/2019 13:56:34	
PB124760BL	S	Phenolics-	P	0.0054	mg/l	11/15/2019 13:56:35	
PB124760BS	S	Phenolics-	P	0.9918	mg/l	11/15/2019 13:56:36	
K5804-01	S	Phenolics-	P	0.0113	mg/l	11/15/2019 13:56:37	
K5804-04	S	Phenolics-	P	0.0049	mg/l	11/15/2019 13:56:38	
K5804-04DUP	S	Phenolics-	P	0.0029	mg/l	11/15/2019 13:56:39	
K5804-04MS	S	Phenolics-	P	0.8626	mg/l	11/15/2019 13:56:40	
K5804-04MSD	S	Phenolics-	P	0.8586	mg/l	11/15/2019 13:56:41	
K5828-01	S	Phenolics-	P	0.0055	mg/l	11/15/2019 13:56:42	
K5828-02	S	Phenolics-	P	0.0045	mg/l	11/15/2019 14:01:53	
K5828-03	S	Phenolics-	P	0.0132	mg/l	11/15/2019 14:01:54	
CCV2	S	Phenolics-	P	0.9595	mg/l	11/15/2019 14:01:55	
CCB2	S	Phenolics-	P	0.0121	mg/l	11/15/2019 14:01:56	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AC

Instrument ID : Konelab

11/15/2019 13:31


Test Phenolics-

Accepted 11/15/2019 13:30

Factor 9.805
Bias 0.006

Coeff. of det. 0.999924

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.011	0.0518	0.0500	
2	Phenol-2	0.016	0.1029	0.1000	
3	Phenol-2	0.032	0.2512	0.2500	
4	Phenol-2	0.057	0.5011	0.5000	
5	Phenol-2	0.107	0.9872	1.0000	
6	Phenol-2	0.211	2.0058	2.0000	