



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 11/19/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 14:00
In Date: 11/18/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:20
Out Date: 11/19/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106354

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K5890-01	MJLLP0	1	1.14	9.98	7.49	71.8	
K5890-02	MJLLP1	2	1.14	9.9	6.46	60.7	
K5890-03	MJLLP2	3	1.14	9.85	7.03	67.6	
K5890-04	MJLLP2D	4	1.14	9.85	7.03	67.6	
K5890-05	MJLLP2S	5	1.14	9.85	7.03	67.6	
K5890-06	MJLLP4	6	1.14	9.56	6.56	64.4	
K5890-07	MJLLP5	7	1.15	9.94	7.4	71.1	
K5890-08	MJLLP6	8	1.16	9.7	6.24	59.5	
K5890-09	MJLLP7	9	1.14	9.52	7.06	70.6	
K5890-10	MJLLP8	10	1.15	9.6	7.05	69.8	
K5890-11	MJLLP9	11	1.14	9.65	7.2	71.2	
K5890-12	MJLLQ0	12	1.16	9.56	6.67	65.6	
K5890-13	MJLLQ2	13	1.12	9.95	6.61	62.2	
K5890-14	MJLLQ3	14	1.12	9.51	6.29	61.6	
K5890-15	MJLLQ4	15	1.12	9.88	6.43	60.6	
K5890-16	MJLLQ5	16	1.13	9.8	6.57	62.7	
K5890-17	MJLLQ6	17	1.13	9.96	7.03	66.8	
K5890-18	MJLLQ7	18	1.13	9.98	6.18	57.1	
K5890-19	MJLLQ8	19	1.13	9.51	5.39	50.8	
K5890-20	MJLLQ9	20	1.12	9.98	7.19	68.5	
K5890-21	MJLLR0	21	1.12	9.56	6.69	66.0	
K5890-22	MJLLR1	22	1.13	9.8	7.14	69.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$