



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB106355

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	11/18/2019	10:25
2	K5822-02		1	9	100.00	>212.0	>212.0	11/18/2019	10:50
3	K5822-02DUP		1	9	100.00	>212.0	>212.0	11/18/2019	11:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB106355

WorkList Name : fp111819 WorkList ID : 133418 Department : Wet-Chemistry Date : 11-18-2019 09:57:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/18/2019	Water	K5822-02 B	Flash Point	Cool 4 deg C	PSEG01	N62	N35868	11/13/2019	1010A

Date/Time 11/18/19 10:00
 Received by: KS
 Relinquished by: DK

Date/Time 11/18/19 11:20
 Received by: DK
 Relinquished by: KS