



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106365

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/18/2019	15:10
2	CAL2	1	Water	NA	NA	20.2	7.00	11/18/2019	15:11
3	CAL3	1	Water	NA	NA	20.2	10.03	11/18/2019	15:12
4	ICV	1	Water	NA	NA	20.2	7.00	11/18/2019	15:15
5	CCV1	1	Water	NA	NA	20.3	2.02	11/18/2019	15:16
6	K5920-01	1	Solid	20.03	20	21.9	8.74	11/18/2019	15:19
7	K5920-01DUP	1	Solid	20.04	20	21.8	8.75	11/18/2019	15:20
8	CCV2	1	Water	NA	NA	20.2	12.02	11/18/2019	15:22

LB106365

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PHS K5920 WorkList ID : 133444 Department : Wet-Chemistry Date : 11-18-2019 14:47:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/18/2019	Solid	K5920-01 - 1	pH	Cool 4 deg C	BRO004	I13	SAMPLE-1	11/18/2019	9045D

Date/Time 11-18-19 15:00
Received by: Jm
Relinquished by: cl

Date/Time 11-18-19 16:00
Received by: Jm
Relinquished by: Jm