



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106366

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/18/2019	15:10
2	CAL2	1	Water	NA	NA	20.2	7.00	11/18/2019	15:11
3	CAL3	1	Water	NA	NA	20.2	10.03	11/18/2019	15:12
4	ICV	1	Water	NA	NA	20.2	7.00	11/18/2019	15:15
5	CCV1	1	Water	NA	NA	20.3	2.01	11/18/2019	15:25
6	K5921-01	1	Water	NA	NA	14.8	7.23	11/18/2019	15:26
7	K5921-01DUP	1	Water	NA	NA	14.9	7.24	11/18/2019	15:27
8	CCV2	1	Water	NA	NA	20.2	12.02	11/18/2019	15:30

2B106366

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH WK5921 WorkList ID : 133445 Department : Wet-Chemistry Date : 11-18-2019 14:48:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/18/2019	Water	K5921-01-17	pH	Cool 4 deg C	DEWB01	N63	WS-1	11/18/2019	9040C

Date/Time 11-18-19 15:00
 Received by: Jm
 Relinquished by: KL

Date/Time 11-18-19 16:10
 Received by: CS
 Relinquished by: Jm