


LB106372

INSTRUMENT ID CVI

7470A



Linear

A= 0.0000e+000

B= 2.3900e-004

C= -6.8955e-002

Rho= 0.9996911

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.069	-0.069	-1	0.000	-1				
0.2	0.200	0.146	-0.054	900	0.0 %	900				
2.5	2.500	2.677	0.177	11488	0.0 %	11488				
5.0	5.000	5.039	0.039	21372	0.0 %	21372				
7.5	7.500	7.400	-0.100	31251	0.0 %	31251				
10.0	10.000	10.007	0.007	42160	0.0 %	42160				

LB106372 INSTRUMENT ID CV1

Method: 7470A	Operator: Admin	Date of Analysis: 19 Nov 2019 07:32:14							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-1	-	0.0000	7470A	PPB	19 Nov 2019 07:38:10	S	Std
0.2 - 1	50.2	900	-	0.2000	7470A	PPB	19 Nov 2019 07:40:14	S	Std
2.5 - 1	52.5	11488	-	2.5000	7470A	PPB	19 Nov 2019 07:42:20	S	Std
5.0 - 1	55.0	21372	-	5.0000	7470A	PPB	19 Nov 2019 07:54:25	S	Std
7.5 - 1	57.5	31251	-	7.5000	7470A	PPB	19 Nov 2019 07:56:32	S	Std
10.0 - 1	510.0	42160	-	10.0000	7470A	PPB	19 Nov 2019 08:11:11	S	Std
ICV10 - 1	ICV10	15870	3.7240	-	7470A	PPB	19 Nov 2019 08:21:14	U	SMPL
ICB10 - 1	ICB10	-53	-0.0816	-	7470A	PPB	19 Nov 2019 08:23:20	U	SMPL
CCV82 - 1	CCV82	21404	5.0466	-	7470A	PPB	19 Nov 2019 08:25:26	U	SMPL
CCB82 - 1	CCB82	-78	-0.0876	-	7470A	PPB	19 Nov 2019 08:27:32	U	SMPL
CRA - 1	CRA	969	0.1626	-	7470A	PPB	19 Nov 2019 08:29:38	U	SMPL
HighStd - 1	HighStd	42585	10.1089	-	7470A	PPB	19 Nov 2019 08:31:43	U	SMPL
ChkStd - 1	ChkStd	31420	7.4405	-	7470A	PPB	19 Nov 2019 08:33:50	U	SMPL
PB124922BL - 1	PBW	-118	-0.0972	-	7470A	PPB	19 Nov 2019 08:35:57	U	SMPL
PB124922BS - 1	LCSW	15433	3.6195	-	7470A	PPB	19 Nov 2019 08:38:05	U	SMPL
K5669-08 - 1	CL-02-111519	197	-0.0219	-	7470A	PPB	19 Nov 2019 08:40:11	U	SMPL
K5669-08DUP - 1	CL-02-111519DUP	271	-0.0042	-	7470A	PPB	19 Nov 2019 08:42:19	U	SMPL
K5669-08MS - 1	CL-02-111519MS	15858	3.7211	-	7470A	PPB	19 Nov 2019 08:44:27	U	SMPL
K5669-08MSD - 1	CL-02-111519MSD	15838	3.7163	-	7470A	PPB	19 Nov 2019 08:46:35	U	SMPL
K5676-20 - 1	SU-01-111519	-63	-0.0840	-	7470A	PPB	19 Nov 2019 08:48:41	U	SMPL
CCV83 - 1	CCV83	21489	5.0669	-	7470A	PPB	19 Nov 2019 08:50:47	U	SMPL
CCB83 - 1	CCB83	-62	-0.0838	-	7470A	PPB	19 Nov 2019 08:52:54	U	SMPL
K5864-11 - 1	SB-01-COMP	4	-0.0680	-	7470A	PPB	19 Nov 2019 08:55:00	U	SMPL
K5864-12 - 1	SB-02-COMP	13	-0.0658	-	7470A	PPB	19 Nov 2019 08:57:07	U	SMPL
K5864-13 - 1	SB-03-COMP	22	-0.0637	-	7470A	PPB	19 Nov 2019 08:59:13	U	SMPL
K5864-14 - 1	SB-04-COMP	29	-0.0620	-	7470A	PPB	19 Nov 2019 09:01:20	U	SMPL
K5864-15 - 1	SB-05-COMP	46	-0.0580	-	7470A	PPB	19 Nov 2019 09:03:27	U	SMPL
K5905-03 - 1	RT-3037	127	-0.0386	-	7470A	PPB	19 Nov 2019 09:05:34	U	SMPL
K5906-03 - 1	NA-SP	70	-0.0522	-	7470A	PPB	19 Nov 2019 09:07:42	U	SMPL
K5669-08LX5 - 1		96	-0.0460	-	7470A	PPB	19 Nov 2019 09:09:49	U	SMPL
PB124922TB - 1	PB124871TB	19	-0.0644	-	7470A	PPB	19 Nov 2019 09:11:57	U	SMPL
CCV84 - 1	CCV84	22300	5.2608	-	7470A	PPB	19 Nov 2019 09:14:04	U	SMPL
CCB84 - 1	CCB84	-30	-0.0761	-	7470A	PPB	19 Nov 2019 09:16:10	U	SMPL