



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB106380

Ambient Barometric Pressure (mmHg): 755.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.9	11/19/2019	11:50
2	K5814-01		1	8	26.00	78.8	79.1	11/19/2019	12:05
3	K5814-01DUP		1	8	27.00	80.6	80.9	11/19/2019	12:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 106380

WorkList Name : fp111919-1 WorkList ID : 133476 Department : Wet-Chemistry Date : 11-19-2019 11:24:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/20/2019	Solid	K5814-01	Flash Point	Cool 4 deg C	PSEG01	N62	20688	11/13/2019	1010A

Date/Time 11/19/19 11:25
 Received by: KS
 Relinquished by: JP

Date/Time 11/19/19 12:30
 Received by: JP
 Relinquished by: KS