



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB106381

Ambient Barometric Pressure (mmHg): 755.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.9	11/20/2019	12:40
2	K5804-01		1	9	100.00	>212.0	>212.0	11/19/2019	13:05
3	K5804-01DUP		1	9	100.00	>212.0	>212.0	11/20/2019	13:32
4	K5804-04		1	8	100.00	>212.0	>212.0	11/20/2019	13:50
5	K5814-02		1	7	100.00	>212.0	>212.0	11/20/2019	14:20
6	K5814-04		1	9	100.00	>212.0	>212.0	11/20/2019	14:40
7	K5814-06		1	9	100.00	>212.0	>212.0	11/20/2019	15:02
8	K5814-08		1	7	100.00	>212.0	>212.0	11/20/2019	15:33
9	K5814-10		1	8	100.00	>212.0	>212.0	11/20/2019	15:58
10	K5828-01		1	9	100.00	>212.0	>212.0	11/20/2019	16:25
11	K5828-02		1	8	100.00	>212.0	>212.0	11/20/2019	16:45
12	K5828-03		1	9	100.00	>212.0	>212.0	11/20/2019	17:10
13	K5852-04		1	7	100.00	>212.0	>212.0	11/20/2019	17:35
14	K5857-03		1	8	100.00	>212.0	>212.0	11/20/2019	17:55
15	K5860-05		1	9	100.00	>212.0	>212.0	11/20/2019	18:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB106381

WorkList Name : fp111919 WorkList ID : 133475 Department : Wet-Chemistry Date : 11-19-2019 11:21:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/25/2019	Water	K5804-01	Flash Point	Cool 4 deg C	PBST01	H51	GE-MW-58-1S	11/12/2019	1010A
11/25/2019	Water	K5804-04	Flash Point	Cool 4 deg C	PBST01	H51	TE-MW-QA-2	11/12/2019	1010A
11/20/2019	Water	K5814-02	Flash Point	Cool 4 deg C	PSEG01	N62	30557	11/13/2019	1010A
11/20/2019	Water	K5814-04	Flash Point	Cool 4 deg C	PSEG01	N62	40189	11/13/2019	1010A
11/20/2019	Water	K5814-06	Flash Point	Cool 4 deg C	PSEG01	N62	30663	11/13/2019	1010A
11/20/2019	Water	K5814-08	Flash Point	Cool 4 deg C	PSEG01	N62	30662	11/13/2019	1010A
11/20/2019	Water	K5814-10	Flash Point	Cool 4 deg C	PSEG01	N62	30535	11/13/2019	1010A
11/26/2019	Water	K5828-01	Flash Point	Cool 4 deg C	PBST01	H51	GE-MW-58-1D	11/13/2019	1010A
11/26/2019	Water	K5828-02	Flash Point	Cool 4 deg C	PBST01	H51	GE-MW-57A-2	11/13/2019	1010A
11/26/2019	Water	K5828-03	Flash Point	Cool 4 deg C	PBST01	H51	REP111319	11/13/2019	1010A
11/21/2019	Water	K5852-04	Flash Point	Cool 4 deg C	PSEG01	N63	20924	11/14/2019	1010A
11/21/2019	Water	K5857-03	Flash Point	Cool 4 deg C	PSEG01	N63	COMP-4-1	11/14/2019	1010A
11/21/2019	Water	K5860-05	Flash Point	Cool 4 deg C	PSEG01	N62	20983-32007	11/14/2019	1010A

Date/Time 11/19/19
 Received by: KS
 Relinquished by: jo

Date/Time 11/19/19 18:25
 Received by: jo
 Relinquished by: KS