



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106390

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/19/2019	15:50
2	CAL2	1	Water	NA	NA	20.1	6.99	11/19/2019	15:51
3	CAL3	1	Water	NA	NA	20.1	10.03	11/19/2019	15:52
4	ICV	1	Water	NA	NA	20.1	7.00	11/19/2019	15:55
5	CCV1	1	Water	NA	NA	20.2	2.01	11/19/2019	16:00
6	K5939-01	1	Water	NA	NA	11.7	12.67	11/19/2019	16:01
7	K5939-01DUP	1	Water	NA	NA	11.8	12.68	11/19/2019	16:02
8	CCV2	1	Water	NA	NA	20.1	12.02	11/19/2019	16:05

LB 106390

WORKLIST(Hardcopy Internal Chain)

WorkList Name :	PH K5939	WorkList ID :	133492	Department :	Wet-Chemistry	Date :	11-19-2019 15:38:31		
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/19/2019	Water	K5939-01 ✓	pH	Cool 4 deg C	JACO05	I21	IDW-DECOM-WATER-111919	11/19/2019	9040C

Date/Time 11-19-19 15:40
Received by: Jn
Relinquished by: JT

Date/Time 11-19-19 16:30
Received by: JT
Relinquished by: JP