



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 11/20/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 11/19/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:50
Out Date: 11/20/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106393

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
K5920-01	SAMPLE-1	1	1.14	9.97	9.2	91.3	
K5927-01	B-1	2	1.18	9.58	8.00	81.2	
K5927-02	B-2	3	1.12	9.77	7.96	79.1	
K5927-03	B-3	4	1.15	9.89	8.29	81.7	
K5927-04	B-4A	5	1.13	9.53	7.95	81.2	
K5927-05	B-4B	6	1.18	9.98	8.29	80.8	
K5927-06	B-5A	7	1.14	9.78	8.15	81.1	
K5927-07	B-5B	8	1.15	9.94	8.49	83.5	
K5927-08	B-6A	9	1.15	9.97	8.65	85.0	
K5927-09	B-7	10	1.1	9.94	8.64	85.3	
K5927-10	B-8	11	1.14	9.92	8.44	83.1	
K5931-01	VNJ-214	12	1.00	2.00	2.00	100.0	STONE SAMPLE 100% SOLIDS
K5932-01	RBR200027-1A-415	13	1.16	9.73	8.6	86.8	
K5933-01	WC-01-1.5-2.0	14	1.12	9.54	8.56	88.4	
K5933-02	WC-01-5.5-6.0	15	1.16	9.67	8.72	88.8	
K5933-03	WC-01-13.5-14.0	16	1.15	9.94	8.33	81.7	
K5933-04	WC-COMP-01	17	1.13	9.54	8.87	92.0	
K5933-07	WC-02-1.5-2.0	18	1.17	9.98	9.11	90.1	
K5933-08	WC-02-5.5-6.0	19	1.12	9.69	8.63	87.6	
K5933-09	WC-02-13.5-14.0	20	1.14	9.96	9.41	93.8	
K5933-10	WC-COMP-02	21	1.17	9.97	9.15	90.7	
K5934-01	TP-1	22	1.14	9.61	7.74	77.9	
K5934-02	TP-2	23	1.16	9.84	7.07	68.1	
K5937-01	RT-1488-RT4221-RBR200031	24	1.17	9.7	8.76	89.0	
K5938-01	ETGI-277	25	1.18	9.93	9.18	91.4	
K5943-01	391112214.1	26	1.16	9.97	1.3	1.6	SLUDGE SAMPLE
K5944-01	391113718.2	27	1.17	9.98	1.24	0.8	SLUDGE SAMPLE
K5946-01	01-A-01-B-01-C	28	1.00	2.00	2.00	100.0	CLUCK SAMPLE 100% SOLIDS

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$