



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106394

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/19/2019	17:25
2	CAL2	1	Water	NA	NA	20.3	7.01	11/19/2019	17:26
3	CAL3	1	Water	NA	NA	20.3	10.03	11/19/2019	17:27
4	ICV	1	Water	NA	NA	20.3	7.00	11/19/2019	17:30
5	CCV1	1	Water	NA	NA	20.2	2.01	11/19/2019	17:32
6	K5916-02	1	Solid	20.03	20	21.8	7.57	11/19/2019	17:34
7	K5933-05	1	Solid	20.02	20	21.2	5.29	11/19/2019	17:35
8	K5933-11	1	Solid	20.03	20	21.4	5.40	11/19/2019	17:36
9	K5934-01	1	Solid	20.04	20	21.6	7.51	11/19/2019	17:38
10	K5934-02	1	Solid	20.01	20	22.0	6.57	11/19/2019	17:40
11	K5938-02	1	Solid	20.03	20	21.7	8.76	11/19/2019	17:42
12	K5938-02DUP	1	Solid	20.04	20	21.8	8.77	11/19/2019	17:43
13	CCV2	1	Water	NA	NA	20.2	12.02	11/19/2019	17:45

106394

WORKLIST(Hardcopy Internal Chain)

WorkList Name: CORROSIONITY K5916

WorkList ID: 133480

Department: Wet-Chemistry

Date: 11-19-2019 12:43:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/18/2019	Solid	K5916-02	Corrosivity	Cool 4 deg C	PSEG01	N63	600429836	11/18/2019	9045D
11/18/2019	Solid	K5933-05	Corrosivity	Cool 4 deg C	PSEG01	N11	WC-COMP-01	11/18/2019	9045D
11/18/2019	Solid	K5933-11	Corrosivity	Cool 4 deg C	PSEG01	N11	WC-COMP-02	11/18/2019	9045D
11/19/2019	Solid	K5934-01	Corrosivity	Cool 4 deg C	PSEG01	N11	TP-1	11/19/2019	9045D
11/19/2019	Solid	K5934-02	Corrosivity	Cool 4 deg C	PSEG01	N11	TP-2	11/19/2019	9045D
11/19/2019	Solid	K5938-02	Corrosivity	Cool 4 deg C	PSEG01	N11	ETGI-277	11/19/2019	9045D

Date/Time 11-19-19 16:10

Received by: Jø

Relinquished by: Jø

Date/Time 11-14-19 18:00

Received by: Jø

Relinquished by: Jø