


LB106403

245.1

INSTRUMENT ID CUT



Linear

A= 0.0000e+000

B= 2.1266e-004

C= -1.3679e-002

Rho= 0.9970745

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.015	-0.015	-7	0.000	-7				
0.05	0.050	0.056	0.006	329	0.0 %	329				
0.2	0.200	0.168	-0.032	856	0.0 %	856				
2.5	2.500	2.323	-0.177	10987	0.0 %	10987				
5.0	5.000	5.210	0.210	24561	0.0 %	24561				
7.5	7.500	7.998	0.498	37671	0.0 %	37671				
10.0	10.000	9.511	-0.489	44786	0.0 %	44786				

LB106403 INSTRUNMRNT ID CV1

Method: 245.1

Operator: Admin

Date of Analysis: 20 Nov 2019 10:59:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-7	-	0.0000	245.1	PPB	20 Nov 2019 14:18:58	S	Std
0.05 - 1	50.05	329	-	0.0500	245.1	PPB	20 Nov 2019 14:21:15	S	Std
0.2 - 1	50.2	856	-	0.2000	245.1	PPB	20 Nov 2019 14:23:32	S	Std
2.5 - 1	52.5	10987	-	2.5000	245.1	PPB	20 Nov 2019 14:25:49	S	Std
5.0 - 1	55.0	24561	-	5.0000	245.1	PPB	20 Nov 2019 14:34:25	S	Std
7.5 - 1	57.5	37671	-	7.5000	245.1	PPB	20 Nov 2019 14:36:40	S	Std
10.0 - 1	57.0	44786	-	10.0000	245.1	PPB	20 Nov 2019 14:50:36	S	Std
ICV12 - 1	ICV12	18844	3.9937	-245.1		PPB	20 Nov 2019 14:59:30	U	SMPL
ICB12 - 1	ICB12	-88	-0.0324	-245.1		PPB	20 Nov 2019 15:01:45	U	SMPL
CCV91 - 1	CCV91	23741	5.0351	-245.1		PPB	20 Nov 2019 15:07:17	U	SMPL
CCB91 - 1	CCB91	-104	-0.0358	-245.1		PPB	20 Nov 2019 15:09:31	U	SMPL
CRA - 1	CRA	363	0.0635	-245.1		PPB	20 Nov 2019 15:11:47	U	SMPL
HighStd - 1	HighStd	51550	10.9491	-245.1		PPB	20 Nov 2019 15:14:02	U	SMPL
ChkStd - 1	ChkStd	31199	6.6212	-245.1		PPB	20 Nov 2019 15:21:14	U	SMPL
PB124985BL - 1	PBW	-95	-0.0339	-245.1		PPB	20 Nov 2019 15:23:30	U	SMPL
PB124985BS - 1	LCSW	16784	3.5556	-245.1		PPB	20 Nov 2019 15:29:13	U	SMPL
K5834-10 - 1	14B-(1-4)-COMP	1216	0.2449	-245.1		PPB	20 Nov 2019 15:31:29	U	SMPL
K5837-02 - 1	EFF-WW	401	0.0716	-245.1		PPB	20 Nov 2019 15:33:46	U	SMPL
K5837-02DUP - 1	EFF-WWDUP	444	0.0807	-245.1		PPB	20 Nov 2019 15:36:42	U	SMPL
K5837-02MS - 1	EFF-WWMS	12147	2.5695	-245.1		PPB	20 Nov 2019 15:39:00	U	SMPL
K5837-02MSD - 1	EFF-WWMSD	12294	2.6008	-245.1		PPB	20 Nov 2019 15:41:15	U	SMPL
CCV92 - 1	CCV92	22391	4.7480	-245.1		PPB	20 Nov 2019 15:43:32	U	SMPL
CCB92 - 1	CCB92	66	0.0004	-245.1		PPB	20 Nov 2019 15:45:50	U	SMPL
K5859-07 - 1	14B-(1-4)-COMP	1822	0.3738	-245.1		PPB	20 Nov 2019 15:48:06	U	SMPL
K5837-02LX5 - 1		42	-0.0047	-245.1		PPB	20 Nov 2019 15:50:21	U	SMPL
K5837-02A - 1		11968	2.5315	-245.1		PPB	20 Nov 2019 15:52:38	U	SMPL
DOC-1 - 1		17685	3.7473	-245.1		PPB	20 Nov 2019 16:00:40	U	SMPL
DOC-2 - 1		17089	3.6205	-245.1		PPB	20 Nov 2019 16:02:56	U	SMPL
DOC-3 - 1		16632	3.5233	-245.1		PPB	20 Nov 2019 16:12:34	U	SMPL
DOC-4 - 1		17128	3.6288	-245.1		PPB	20 Nov 2019 16:14:50	U	SMPL
CCV93 - 1	CCV93	22727	4.8195	-245.1		PPB	20 Nov 2019 16:17:07	U	SMPL
CCB93 - 1	CCB93	264	0.0425	-245.1		PPB	20 Nov 2019 16:19:25	U	SMPL