



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106407

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/20/2019	17:10
2	CAL2	1	Water	NA	NA	20.1	7.00	11/20/2019	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	11/20/2019	17:12
4	ICV	1	Water	NA	NA	20.1	6.99	11/20/2019	17:15
5	CCV1	1	Water	NA	NA	20.1	2.01	11/20/2019	17:16
6	K5676-25	1	Solid	20.03	20	21.2	7.60	11/20/2019	17:18
7	K5676-25DUP	1	Solid	20.04	20	21.3	7.61	11/20/2019	17:19
8	K5676-30	1	Solid	20.05	20	21.4	9.60	11/20/2019	17:20
9	K5948-03	1	Solid	20.03	20	21.2	9.46	11/20/2019	17:22
10	CCV2	1	Water	NA	NA	21.3	12.02	11/20/2019	17:25

28106407

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K5676 WorkList ID : 133528 Department : Wet-Chemistry Date : 11-20-2019 15:36:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/20/2019	Solid	K5676-25	Corrosivity	Cool 4 deg C	PSEG05		SU-02-112019	11/20/2019	9045D
11/20/2019	Solid	K5676-30	Corrosivity	Cool 4 deg C	PSEG05		SU-03-112019	11/20/2019	9045D
11/19/2019	Solid	K5948-03	Corrosivity	Cool 4 deg C	WSP001	I22	BSW-1-0-12	11/19/2019	9045D

Date/Time 11-20-19 16:15
 Received by: Jo
 Relinquished by: J.C

Date/Time 11-20-19 18:00
 Received by: J.C
 Relinquished by: J.C