


LB106416
INSTRUMENT ID CUI

7470A

μ Abs.:

45784



Linear

A= 0.0000e+000

B= 2.1979e-004

C= 7.8519e-003

Rho= 0.9998543

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.006	0.006	-9	0.000	-9				
0.2	0.200	0.216	0.016	945	0.0 %	945				
2.5	2.500	2.484	-0.016	11266	0.0 %	11266				
5.0	5.000	5.049	0.049	22937	0.0 %	22937				
7.5	7.500	7.374	-0.126	33515	0.0 %	33515				
10.0	10.000	10.071	0.071	45784	0.0 %	45784				

LB106416

Method: 7470A

Operator: Admin

Date of Analysis: 21 Nov 2019 08:26:03

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	5.0.0	-9	-	0.0000	7470A	PPB	21 Nov 2019 09:10:06	S	Std
0.2 - 1	5.0.2	945	-	0.2000	7470A	PPB	21 Nov 2019 09:12:14	S	Std
2.5 - 1	5.2.5	0	-	2.5000	7470A	PPB	21 Nov 2019 09:14:20	S	Std
5.0 - 1	5.5.0	22937	-	5.0000	7470A	PPB	21 Nov 2019 09:16:54	S	Std
7.5 - 1	5.7.5	33515	-	7.5000	7470A	PPB	21 Nov 2019 09:19:01	S	Std
10.0 - 1	5.10.0	45784	-	10.0000	7470A	PPB	21 Nov 2019 09:23:29	S	Std
ICV13 - 1	ICV13	17515	3.8576	-	7470A	PPB	21 Nov 2019 09:31:28	U	SMPL
ICB13 - 1	ICB13	62	0.0215	-	7470A	PPB	21 Nov 2019 09:33:34	U	SMPL
CCV01 - 1	CCV01	22908	5.0429	-	7470A	PPB	21 Nov 2019 09:35:41	U	SMPL
CCB01 - 1	CCB01	21	0.0125	-	7470A	PPB	21 Nov 2019 09:37:47	U	SMPL
CRA - 1	CRA	962	0.2193	-	7470A	PPB	21 Nov 2019 09:39:53	U	SMPL
HighStd - 1	HighStd	46138	10.1487	-	7470A	PPB	21 Nov 2019 09:41:59	U	SMPL
ChkStd - 1	ChkStd	31548	6.9419	-	7470A	PPB	21 Nov 2019 09:44:05	U	SMPL
PB124995BL - 1	PBW	8	0.0096	-	7470A	PPB	21 Nov 2019 09:46:12	U	SMPL
PB124995BS - 1	LCSW	17323	3.8154	-	7470A	PPB	21 Nov 2019 09:48:19	U	SMPL
K5916-02 - 1	600429836	80	0.0254	-	7470A	PPB	21 Nov 2019 09:50:26	U	SMPL
K5916-02DUP - 1	600429836DUP	95	0.0287	-	7470A	PPB	21 Nov 2019 09:52:34	U	SMPL
K5916-02MS - 1	600429836MS	18375	4.0466	-	7470A	PPB	21 Nov 2019 09:54:42	U	SMPL
K5916-02MSD - 1	600429836MSD	18383	4.0483	-	7470A	PPB	21 Nov 2019 10:00:04	U	SMPL
K5923-03 - 1	SP-1	-44	-0.0018	-	7470A	PPB	21 Nov 2019 10:02:10	U	SMPL
CCV02 - 1	CCV02	22664	4.9893	-	7470A	PPB	21 Nov 2019 10:04:14	U	SMPL
CCB02 - 1	CCB02	6	0.0092	-	7470A	PPB	21 Nov 2019 10:06:20	U	SMPL
K5933-05 - 1	WC-COMP-01	-6	0.0065	-	7470A	PPB	21 Nov 2019 10:08:27	U	SMPL
K5933-11 - 1	WC-COMP-02	44	0.0175	-	7470A	PPB	21 Nov 2019 10:10:31	U	SMPL
K5934-01 - 1	TP-1	55	0.0199	-	7470A	PPB	21 Nov 2019 10:12:38	U	SMPL
K5934-02 - 1	TP-2	76	0.0246	-	7470A	PPB	21 Nov 2019 10:14:44	U	SMPL
K5938-02 - 1	ETGI-277	28	0.0140	-	7470A	PPB	21 Nov 2019 10:19:02	U	SMPL
K5916-02LX5 - 1		3	0.0085	-	7470A	PPB	21 Nov 2019 10:21:09	U	SMPL
PB124995TB - 1	PB124957TB	42	0.0171	-	7470A	PPB	21 Nov 2019 10:23:16	U	SMPL
CCV03 - 1	CCV03	22507	4.9548	-	7470A	PPB	21 Nov 2019 10:28:06	U	SMPL
CCB03 - 1	CCB03	40	0.0166	-	7470A	PPB	21 Nov 2019 10:30:14	U	SMPL
DOC-1 - 1		17226	3.7940	-	7470A	PPB	21 Nov 2019 10:32:20	U	SMPL
DOC-2 - 1		18024	3.9694	-	7470A	PPB	21 Nov 2019 10:34:26	U	SMPL
DOC-3 - 1		16144	3.5562	-	7470A	PPB	21 Nov 2019 10:36:33	U	SMPL
DOC-4 - 1		15164	3.3408	-	7470A	PPB	21 Nov 2019 10:38:39	U	SMPL
CCV04 - 1	CCV04	22532	4.9603	-	7470A	PPB	21 Nov 2019 10:40:45	U	SMPL
CCB04 - 1	CCB04	17	0.0116	-	7470A	PPB	21 Nov 2019 10:42:52	U	SMPL