

LB106420


7471B

INSTRUMENT ID CVI

Linear

 μ Abs.:

4859E



A= 0.0000e+000

B= 2.0495e-004

C= -4.8912e-002

Rho= 0.9998637

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.042	-0.042	36	0.000	36				
0.2	0.200	0.156	-0.044	999	0.0 %	999				
2.5	2.500	2.565	0.065	12756	0.0 %	12756				
5.0	5.000	5.042	0.042	24841	0.0 %	24841				
7.5	7.500	7.567	0.067	37162	0.0 %	37162				
10.0	10.000	9.911	-0.089	48595	0.0 %	48595				

LB106420 INSTRUMENT ID CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Nov 2019 11:39:19

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50.0	36	-	0.00007471B	PPB	21 Nov 2019 11:44:23	S	Std
0.2 - 1	50.2	999	-	0.20007471B	PPB	21 Nov 2019 11:46:40	S	Std
2.5 - 1	52.5	12756	-	2.50007471B	PPB	21 Nov 2019 11:48:57	S	Std
5.0 - 1	55.0	24841	-	5.00007471B	PPB	21 Nov 2019 11:51:14	S	Std
7.5 - 1	57.5	37162	-	7.50007471B	PPB	21 Nov 2019 11:53:31	S	Std
10.0 - 1	510.0	48595	-	10.00007471B	PPB	21 Nov 2019 12:14:06	S	Std
ICV14 - 1	ICV14	20866	4.2276	-7471B	PPB	21 Nov 2019 12:49:03	U	SMPL
ICB14 - 1	ICB14	-16	-0.0522	-7471B	PPB	21 Nov 2019 12:51:18	U	SMPL
CCV05 - 1	CCV05	23318	4.7301	-7471B	PPB	21 Nov 2019 12:59:10	U	SMPL
CCB05 - 1	CCB05	61	-0.0364	-7471B	PPB	21 Nov 2019 13:03:03	U	SMPL
CRA - 1	CRA	1085	0.1735	-7471B	PPB	21 Nov 2019 13:05:18	U	SMPL
HighStd - 1	HighStd	54461	11.1128	-7471B	PPB	21 Nov 2019 13:07:33	U	SMPL
ChkStd - 1	ChkStd	34099	6.9397	-7471B	PPB	21 Nov 2019 13:09:49	U	SMPL
PB124986BL - 1	PBS	-6	-0.0501	-7471B	PPB	21 Nov 2019 13:12:07	U	SMPL
PB124986BS - 1	LCSS	18849	3.8142	-7471B	PPB	21 Nov 2019 13:14:25	U	SMPL
K5916-01 - 1	600429836	7044	1.3948	-7471B	PPB	21 Nov 2019 13:16:42	U	SMPL
K5916-01DUP - 1	600429836DUP	6961	1.3777	-7471B	PPB	21 Nov 2019 13:19:00	U	SMPL
K5916-01MS - 1	600429836MS	24495	4.9713	-7471B	PPB	21 Nov 2019 13:21:18	U	SMPL
K5916-01MSD - 1	600429836MSD	22726	4.6088	-7471B	PPB	21 Nov 2019 13:23:35	U	SMPL
K5918-01 - 1	SOIL-DRUM	1386	0.2351	-7471B	PPB	21 Nov 2019 13:25:52	U	SMPL
CCV06 - 1	CCV06	23128	4.6912	-7471B	PPB	21 Nov 2019 13:28:09	U	SMPL
CCB06 - 1	CCB06	-36	-0.0563	-7471B	PPB	21 Nov 2019 13:30:25	U	SMPL
K5925-01 - 1	CARRIAGE-COURT	1024	0.1610	-7471B	PPB	21 Nov 2019 13:32:42	U	SMPL
K5927-01 - 1	B-1	222	-0.0034	-7471B	PPB	21 Nov 2019 13:34:57	U	SMPL
K5927-02 - 1	B-2	219	-0.0040	-7471B	PPB	21 Nov 2019 13:37:13	U	SMPL
K5927-03 - 1	B-3	1475	0.2534	-7471B	PPB	21 Nov 2019 13:41:34	U	SMPL
K5927-04 - 1	B-4A	830	0.1212	-7471B	PPB	21 Nov 2019 13:43:49	U	SMPL
K5927-05 - 1	B-4B	2552	0.4741	-7471B	PPB	21 Nov 2019 13:46:06	U	SMPL
K5927-06 - 1	B-5A	2104	0.3823	-7471B	PPB	21 Nov 2019 13:48:23	U	SMPL
K5927-07 - 1	B-5B	446	0.0425	-7471B	PPB	21 Nov 2019 13:50:39	U	SMPL
K5927-08 - 1	B-6A	1433	0.2448	-7471B	PPB	21 Nov 2019 13:52:57	U	SMPL
K5927-09 - 1	B-7	247	0.0017	-7471B	PPB	21 Nov 2019 13:55:13	U	SMPL
CCV07 - 1	CCV07	22881	4.6405	-7471B	PPB	21 Nov 2019 13:57:28	U	SMPL
CCB07 - 1	CCB07	-61	-0.0614	-7471B	PPB	21 Nov 2019 13:59:44	U	SMPL
K5927-10 - 1	B-8	1335	0.2247	-7471B	PPB	21 Nov 2019 14:02:01	U	SMPL
K5932-01 - 1	RBR200027-1A-415	3375	0.6428	-7471B	PPB	21 Nov 2019 14:04:16	U	SMPL
K5933-04 - 1	WC-COMP-01	403	0.0337	-7471B	PPB	21 Nov 2019 14:06:31	U	SMPL
K5933-10 - 1	WC-COMP-02	389	0.0308	-7471B	PPB	21 Nov 2019 14:08:47	U	SMPL
K5937-01 - 1	RT-1488-RT4221-RBR200031	1820	0.3241	-7471B	PPB	21 Nov 2019 14:11:03	U	SMPL
K5938-01 - 1	ETGI-277	2663	0.4969	-7471B	PPB	21 Nov 2019 14:13:19	U	SMPL
CCV08 - 1	CCV08	23766	4.8219	-7471B	PPB	21 Nov 2019 14:17:53	U	SMPL
CCB08 - 1	CCB08	-45	-0.0581	-7471B	PPB	21 Nov 2019 14:20:10	U	SMPL
DOC-1 - 1		17724	3.5836	-7471B	PPB	21 Nov 2019 14:22:28	U	SMPL
DOC-2 - 1		17648	3.5680	-7471B	PPB	21 Nov 2019 14:24:44	U	SMPL
DOC-3 - 1		19442	3.9357	-7471B	PPB	21 Nov 2019 14:27:01	U	SMPL
DOC-4 - 1		20741	4.2019	-7471B	PPB	21 Nov 2019 14:29:19	U	SMPL
K5916-01LX5 - 1		1239	0.2050	-7471B	PPB	21 Nov 2019 14:31:36	U	SMPL
CCV09 - 1	CCV09	22992	4.6633	-7471B	PPB	21 Nov 2019 14:33:53	U	SMPL
CCB09 - 1	CCB09	-29	-0.0549	-7471B	PPB	21 Nov 2019 14:36:09	U	SMPL