



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 11/22/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 11/21/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:05
Out Date: 11/22/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106426

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
K5668-08	HD-01-112119	1	1.15	9.5	8.55	88.6	
K5668-09	HD-01-112119	2	1.14	9.66	8.91	91.2	
K5668-10	HD-01-112119-1	3	1.17	9.83	8.78	87.9	
K5670-09	OK-01-112119	4	1.16	9.64	8.61	87.9	
K5670-10	OK-01-112119	5	1.12	9.53	8.34	85.9	
K5670-11	OK-01-112119-1	6	1.16	9.71	8.75	88.8	
K5671-11	OR-02-112119	7	1.17	9.87	8.77	87.4	
K5671-12	OR-02-112119	8	1.14	9.95	8.93	88.4	
K5671-13	OR-02-112119-1	9	1.13	9.97	8.79	86.7	
K5671-14	OR-02-112119-2	10	1.14	9.6	8.48	86.8	
K5966-01	PCB-110419-01	11	1.00	2.00	2.00	100.0	CLUCK SAMPLE 100% SOLIDS
K5966-02	PCB-110519-02	12	1.00	2.00	2.00	100.0	CLUCK SAMPLE 100% SOLIDS
K5966-03	PCB-110619-03	13	1.00	2.00	2.00	100.0	CLUCK SAMPLE 100% SOLIDS
K5966-04	PCB-110619-04	14	1.00	2.00	2.00	100.0	CLUCK SAMPLE 100% SOLIDS
K5966-05	PCB-110619-05	15	1.00	2.00	2.00	100.0	CLUCK SAMPLE 100% SOLIDS
K5966-06	PCB-110719-06	16	1.00	2.00	2.00	100.0	CLUCK SAMPLE 100% SOLIDS
K5966-07	PCB-110719-07	17	1.00	2.00	2.00	100.0	CLUCK SAMPLE 100% SOLIDS
K5970-01	TP-28-S1	18	1.15	9.74	8.56	86.3	
K5970-02	TP-28-S2	19	1.15	9.55	8.56	88.2	
K5970-03	TP-28-S	20	1.14	9.83	8.83	88.5	
K5970-05	TP-28-D1	21	1.18	9.96	8.87	87.6	
K5970-06	TP-28-D2	22	1.14	9.8	8.86	89.1	
K5970-07	TP-28-D	23	1.19	9.74	8.75	88.4	
K5972-01	25441-26560	24	1.18	9.79	8.3	82.7	
K5972-02	25441-26560	25	1.17	9.9	8.48	83.7	
K5972-04	SOIL-PILE	26	1.13	9.75	8.24	82.5	
K5972-05	SOIL-PILE	27	1.13	9.75	8.24	82.5	
K5972-06	SOIL-PILE	28	1.13	9.75	8.24	82.5	



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 11/22/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 11/21/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:05
Out Date: 11/22/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106426

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
K5973-01	DUCT-BANK-1	29	1.15	9.97	9.00	89.0	
K5973-02	DUCT-BANK-1	30	1.17	9.7	8.55	86.5	
K5973-03	DUCT-BANK-1	31	1.15	9.97	9.00	89.0	
K5973-04	DUCT-BANK-2	32	1.12	9.52	8.21	84.4	
K5973-05	DUCT-BANK-2	33	1.19	9.87	8.19	80.6	
K5973-06	DUCT-BANK-2	34	1.12	9.52	8.21	84.4	
K5973-07	TP-7	35	1.11	9.83	8.48	84.5	
K5973-08	TP-7	36	1.17	9.7	8.44	85.2	
K5973-09	TP-7	37	1.11	9.83	8.48	84.5	
K5977-01	WS-112119-0-1-GRAB-B1	38	1.18	9.76	9.21	93.6	
K5977-02	WS-112119-0-2-COMP-B1	39	1.18	9.68	8.95	91.4	
K5977-03	WS-112119-2-5-COMP-B1	40	1.13	9.63	8.98	92.4	
K5977-04	WS-112119-4-5-GRAB-B1	41	1.18	9.55	8.84	91.5	
K5977-05	WS-112119-0-1-GRAB-B5	42	1.15	9.94	9.33	93.1	
K5977-06	WS-112119-0-2-COMP-B5	43	1.16	9.89	9.29	93.1	
K5977-07	WS-112119-2-5-COMP-B5	44	1.17	9.69	8.95	91.3	
K5977-08	WS-112119-4-5-GRAB-B5	45	1.14	9.64	8.89	91.2	
K5977-09	WS-112119-0-1-GRAB-B6	46	1.13	9.92	8.62	85.2	
K5977-10	WS-112119-0-2-COMP-B6	47	1.15	9.68	8.44	85.5	
K5977-11	WS-112119-2-5-COMP-B6	48	1.18	9.92	8.72	86.3	
K5977-12	WS-112119-4-5-GRAB-B6	49	1.14	9.66	8.46	85.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-112119 WorkList ID : 133530 Department : Wet-Chemistry Date : 11-21-2019 07:46:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/26/2019	Solid	K5668-08	Percent Solids	Cool 4 deg C	PSEG05		HD-01-112119	11/21/2019	Chemtech -SO
11/26/2019	Solid	K5668-09	Percent Solids	Cool 4 deg C	PSEG05		HD-01-112119	11/21/2019	Chemtech -SO
11/26/2019	Solid	K5668-10	Percent Solids	Cool 4 deg C	PSEG05		HD-01-112119-1	11/21/2019	Chemtech -SO
11/26/2019	Solid	K5670-09	Percent Solids	Cool 4 deg C	PSEG05		OK-01-112119	11/21/2019	Chemtech -SO
11/26/2019	Solid	K5670-10	Percent Solids	Cool 4 deg C	PSEG05		OK-01-112119	11/21/2019	Chemtech -SO
11/26/2019	Solid	K5670-11	Percent Solids	Cool 4 deg C	PSEG05		OK-01-112119-1	11/21/2019	Chemtech -SO
11/26/2019	Solid	K5671-11	Percent Solids	Cool 4 deg C	PSEG05		OR-02-112119	11/21/2019	Chemtech -SO
11/26/2019	Solid	K5671-12	Percent Solids	Cool 4 deg C	PSEG05		OR-02-112119	11/21/2019	Chemtech -SO
11/26/2019	Solid	K5671-13	Percent Solids	Cool 4 deg C	PSEG05		OR-02-112119-1	11/21/2019	Chemtech -SO
11/26/2019	Solid	K5671-14	Percent Solids	Cool 4 deg C	PSEG05		OR-02-112119-2	11/21/2019	Chemtech -SO
11/14/2019	Solid	K5966-01	Percent Solids	Cool 4 deg C	CLOU03	I31	PCB-110419-01	11/04/2019	Chemtech -SO
11/15/2019	Solid	K5966-02	Percent Solids	Cool 4 deg C	CLOU03	I31	PCB-110519-02	11/05/2019	Chemtech -SO
11/16/2019	Solid	K5966-03	Percent Solids	Cool 4 deg C	CLOU03	I31	PCB-110619-03	11/06/2019	Chemtech -SO
11/16/2019	Solid	K5966-04	Percent Solids	Cool 4 deg C	CLOU03	I31	PCB-110619-04	11/06/2019	Chemtech -SO
11/16/2019	Solid	K5966-05	Percent Solids	Cool 4 deg C	CLOU03	I31	PCB-110619-05	11/06/2019	Chemtech -SO
11/17/2019	Solid	K5966-06	Percent Solids	Cool 4 deg C	CLOU03	I31	PCB-110719-06	11/07/2019	Chemtech -SO
11/17/2019	Solid	K5966-07	Percent Solids	Cool 4 deg C	CLOU03	I31	PCB-110719-07	11/07/2019	Chemtech -SO
11/28/2019	Solid	K5970-01	Percent Solids	Cool 4 deg C	PSEG01	N11	TP-28-S1	11/21/2019	Chemtech -SO
11/28/2019	Solid	K5970-02	Percent Solids	Cool 4 deg C	PSEG01	N11	TP-28-S2	11/21/2019	Chemtech -SO
11/28/2019	Solid	K5970-03	Percent Solids	Cool 4 deg C	PSEG01	N11	TP-28-S	11/21/2019	Chemtech -SO
11/28/2019	Solid	K5970-05	Percent Solids	Cool 4 deg C	PSEG01	N11	TP-28-D1	11/21/2019	Chemtech -SO
11/28/2019	Solid	K5970-06	Percent Solids	Cool 4 deg C	PSEG01	N11	TP-28-D2	11/21/2019	Chemtech -SO

Date/Time 11-21-19 16:00
 Received by: Jn
 Relinquished by: Jn

Date/Time 11-21-19 17:30
 Received by: Jn
 Relinquished by: Jn

HP106426

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-112119 WorkList ID : 133530 Department : Wet-Chemistry Date : 11-21-2019 07:46:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/28/2019	Solid	K5970-07	Percent Solids	Cool 4 deg C	PSEG01	N11	TP-28-D	11/21/2019	Chemtech -SO
11/27/2019	Solid	K5972-01	Percent Solids	Cool 4 deg C	PSEG01	N11	25441-26560	11/21/2019	Chemtech -SO
11/27/2019	Solid	K5972-02	Percent Solids	Cool 4 deg C	PSEG01	N11	25441-26560	11/21/2019	Chemtech -SO
11/27/2019	Solid	K5972-04	Percent Solids	Cool 4 deg C	PSEG01	N11	SOIL-PILE	11/21/2019	Chemtech -SO
11/27/2019	Solid	K5972-05	Percent Solids	Cool 4 deg C	PSEG01	N11	SOIL-PILE	11/21/2019	Chemtech -SO
11/27/2019	Solid	K5972-06	Percent Solids	Cool 4 deg C	PSEG01	N11	SOIL-PILE	11/21/2019	Chemtech -SO
11/27/2019	Solid	K5973-01	Percent Solids	Cool 4 deg C	PSEG01	N11	DUCT-BANK-1	11/21/2019	Chemtech -SO
11/27/2019	Solid	K5973-02	Percent Solids	Cool 4 deg C	PSEG01	N11	DUCT-BANK-1	11/21/2019	Chemtech -SO
11/27/2019	Solid	K5973-03	Percent Solids	Cool 4 deg C	PSEG01	N11	DUCT-BANK-1	11/21/2019	Chemtech -SO
11/27/2019	Solid	K5973-04	Percent Solids	Cool 4 deg C	PSEG01	N11	DUCT-BANK-2	11/21/2019	Chemtech -SO
11/27/2019	Solid	K5973-05	Percent Solids	Cool 4 deg C	PSEG01	N11	DUCT-BANK-2	11/21/2019	Chemtech -SO
11/27/2019	Solid	K5973-06	Percent Solids	Cool 4 deg C	PSEG01	N11	DUCT-BANK-2	11/21/2019	Chemtech -SO
11/27/2019	Solid	K5973-07	Percent Solids	Cool 4 deg C	PSEG01	N11	TP-7	11/21/2019	Chemtech -SO
11/27/2019	Solid	K5973-08	Percent Solids	Cool 4 deg C	PSEG01	N11	TP-7	11/21/2019	Chemtech -SO
11/27/2019	Solid	K5973-09	Percent Solids	Cool 4 deg C	PSEG01	N11	TP-7	11/21/2019	Chemtech -SO
12/01/2019	Solid	K5977-01	Percent Solids	Cool 4 deg C	ENTE01	I23	WS-112119-0-1-GRAB-B1	11/21/2019	Chemtech -SO
12/01/2019	Solid	K5977-02	Percent Solids	Cool 4 deg C	ENTE01	I23	WS-112119-0-2-COMP-B1	11/21/2019	Chemtech -SO
12/01/2019	Solid	K5977-03	Percent Solids	Cool 4 deg C	ENTE01	I23	WS-112119-2-5-COMP-B1	11/21/2019	Chemtech -SO
12/01/2019	Solid	K5977-04	Percent Solids	Cool 4 deg C	ENTE01	I23	WS-112119-4-5-GRAB-B1	11/21/2019	Chemtech -SO
12/01/2019	Solid	K5977-05	Percent Solids	Cool 4 deg C	ENTE01	I23	WS-112119-0-1-GRAB-B5	11/21/2019	Chemtech -SO
12/01/2019	Solid	K5977-06	Percent Solids	Cool 4 deg C	ENTE01	I23	WS-112119-0-2-COMP-B5	11/21/2019	Chemtech -SO
12/01/2019	Solid	K5977-07	Percent Solids	Cool 4 deg C	ENTE01	I23	WS-112119-2-5-COMP-B5	11/21/2019	Chemtech -SO

Date/Time	11-21-19 16:00	Date/Time	11-21-19 17:30
Received by:	Jo	Received by:	CS
Relinquished by:	CS	Relinquished by:	Jo

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-112119 WorkList ID : 133530 Department : Wet-Chemistry Date : 11-21-2019 07:46:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/01/2019	Solid	K5977-08	Percent Solids	Cool 4 deg C	ENTE01	I23	WS-112119-4-5-GRAB-B5	11/21/2019	Chemtech -SO
12/01/2019	Solid	K5977-09	Percent Solids	Cool 4 deg C	ENTE01	I23	WS-112119-0-1-GRAB-B6	11/21/2019	Chemtech -SO
12/01/2019	Solid	K5977-10	Percent Solids	Cool 4 deg C	ENTE01	I23	WS-112119-0-2-COMP-B6	11/21/2019	Chemtech -SO
12/01/2019	Solid	K5977-11	Percent Solids	Cool 4 deg C	ENTE01	I23	WS-112119-2-5-COMP-B6	11/21/2019	Chemtech -SO
12/01/2019	Solid	K5977-12	Percent Solids	Cool 4 deg C	ENTE01	I23	WS-112119-4-5-GRAB-B6	11/21/2019	Chemtech -SO

Date/Time 11-21-19 16:00
 Received by: JN
 Relinquished by: JN

Date/Time 11-21-19 17:30
 Received by: JN
 Relinquished by: JN