


LB106428
INSTRUMENT ID (V)

ISM02.4



Linear

A= 0.0000e+000

B= 2.8170e-004 *slope*

C= -5.5060e-002 *y-intercept*

Rho= 0.9964944

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.057	-0.057	-7	0.000	-7				0.1
0.05	0.050									
0.20	0.200	0.156	-0.044	748	0.0 %	748				-22
2.50	2.500	2.631	0.131	9534	0.0 %	9534				5
5.00	5.000	4.832	-0.168	17348	0.0 %	17348				-3
7.50	7.500	8.075	0.575	28862	0.0 %	28862				8
10.0	10.000	9.564	-0.436	34146	0.0 %	34146				-4

11/22/19

LB106428

Instrument ID : CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0.0	-7 -		0	ISM02.4	PPB	11/22/2019 7:41	Std
0.2 S0.2	748 -		0.2	ISM02.4	PPB	11/22/2019 7:44	Std
2.5 S2.5	9534 -		2.5	ISM02.4	PPB	11/22/2019 7:46	Std
5 S5.0	17348 -		5	ISM02.4	PPB	11/22/2019 7:48	Std
7.5 S7.5	28862 -		7.5	ISM02.4	PPB	11/22/2019 7:50	Std
10 S10.0	34146 -		10	ISM02.4	PPB	11/22/2019 8:02	Std
ICV	ICV	15513	4.3149 -	ISM02.4	PPB	11/22/2019 8:05	SMPL
ICB	ICB	-32	-0.0641 -	ISM02.4	PPB	11/22/2019 8:08	SMPL
CCV010	CCV010	17782	4.9541 -	ISM02.4	PPB	11/22/2019 8:10	SMPL
CCB010	CCB010	-59	-0.0717 -	ISM02.4	PPB	11/22/2019 8:12	SMPL
PB124990E	PBS990	-5	-0.0565 -	ISM02.4	PPB	11/22/2019 8:14	SMPL
K5892-01	MJLLQ1	560	0.1027 -	ISM02.4	PPB	11/22/2019 8:17	SMPL
K5892-02	MJLLQ1D	569	0.1052 -	ISM02.4	PPB	11/22/2019 8:19	SMPL
K5892-03	MJLLQ1S	9315	2.5689 -	ISM02.4	PPB	11/22/2019 8:25	SMPL
K5892-04	MJLLR2	868	0.1895 -	ISM02.4	PPB	11/22/2019 8:27	SMPL
K5892-05	MJLLR3	253	0.0162 -	ISM02.4	PPB	11/22/2019 8:29	SMPL
K5892-06	MJLLR4	602	0.1145 -	ISM02.4	PPB	11/22/2019 8:31	SMPL
K5892-07	MJLLR5	585	0.1097 -	ISM02.4	PPB	11/22/2019 8:34	SMPL
K5892-08	MJLLR6	422	0.0638 -	ISM02.4	PPB	11/22/2019 8:36	SMPL
K5892-09	MJLLR7	298	0.0289 -	ISM02.4	PPB	11/22/2019 8:38	SMPL
K5892-10	MJLLR8	345	0.0421 -	ISM02.4	PPB	11/22/2019 8:41	SMPL
K5892-11	MJLLR9	336	0.0396 -	ISM02.4	PPB	11/22/2019 8:43	SMPL
K5892-12	MJLLS0	868	0.1895 -	ISM02.4	PPB	11/22/2019 8:45	SMPL
K5892-13	MJLLS1	670	0.1337 -	ISM02.4	PPB	11/22/2019 8:47	SMPL
K5892-14	MJLLS2	821	0.1762 -	ISM02.4	PPB	11/22/2019 8:50	SMPL
K5892-15	MJLLS6	654	0.1292 -	ISM02.4	PPB	11/22/2019 8:52	SMPL
K5892-16	MJLLS7	408	0.0599 -	ISM02.4	PPB	11/22/2019 8:54	SMPL
K5892-17	MJLLS8	496	0.0847 -	ISM02.4	PPB	11/22/2019 8:56	SMPL
PB124991E	PBS991	6	-0.0534 -	ISM02.4	PPB	11/22/2019 8:59	SMPL
K5912-01	MJLNR0	940	0.2097 -	ISM02.4	PPB	11/22/2019 9:01	SMPL
CCV011	CCV011	17307	4.8202 -	ISM02.4	PPB	11/22/2019 9:03	SMPL
CCB011	CCB011	-43	-0.0672 -	ISM02.4	PPB	11/22/2019 9:06	SMPL
K5912-02	MJLNR2	605	0.1154 -	ISM02.4	PPB	11/22/2019 9:08	SMPL
K5912-03	MJLNR3	627	0.1216 -	ISM02.4	PPB	11/22/2019 9:10	SMPL
K5912-04	MJLNR4	286	0.0255 -	ISM02.4	PPB	11/22/2019 9:12	SMPL
K5912-05	MJLNR4D	294	0.0278 -	ISM02.4	PPB	11/22/2019 9:15	SMPL
K5912-06	MJLNR4S	10874	3.0081 -	ISM02.4	PPB	11/22/2019 9:17	SMPL
K5912-07	MJLNR5	285	0.0252 -	ISM02.4	PPB	11/22/2019 9:19	SMPL
K5912-08	MJLNR6	586	0.11 -	ISM02.4	PPB	11/22/2019 9:21	SMPL
K5912-09	MJLNR7	904	0.1996 -	ISM02.4	PPB	11/22/2019 9:24	SMPL
K5912-10	MJLNR8	791	0.1678 -	ISM02.4	PPB	11/22/2019 9:26	SMPL
K5912-11	MJLNR9	883	0.1937 -	ISM02.4	PPB	11/22/2019 9:28	SMPL
K5912-12	MJLNS0	559	0.1024 -	ISM02.4	PPB	11/22/2019 9:30	SMPL
K5912-13	MJLNS1	873	0.1909 -	ISM02.4	PPB	11/22/2019 9:33	SMPL
K5912-14	MJLNS2	825	0.1773 -	ISM02.4	PPB	11/22/2019 9:35	SMPL
K5912-15	MJLNS6	650	0.128 -	ISM02.4	PPB	11/22/2019 9:37	SMPL

LB106428

Instrument ID : CV1

K5912-16	MJLNS7	782	0.1652 -	ISM02.4	PPB	11/22/2019 9:39 SMPL
K5912-17	MJLNS8	786	0.1664 -	ISM02.4	PPB	11/22/2019 9:42 SMPL
K5912-18	MJLNS9	742	0.154 -	ISM02.4	PPB	11/22/2019 9:44 SMPL
K5912-19	MJLNT0	539	0.0968 -	ISM02.4	PPB	11/22/2019 9:46 SMPL
K5912-20	MJLNT1	669	0.1334 -	ISM02.4	PPB	11/22/2019 9:49 SMPL
K5912-21	MJLNT2	659	0.1306 -	ISM02.4	PPB	11/22/2019 9:51 SMPL
CCV012	CCV012	17261	4.8073 -	ISM02.4	PPB	11/22/2019 9:53 SMPL
CCB012	CCB012	36	-0.0449 -	ISM02.4	PPB	11/22/2019 9:55 SMPL
K5912-22	MJLNT3	1450	0.3534 -	ISM02.4	PPB	11/22/2019 9:58 SMPL
PB124992E	PBS992	63	-0.0373 -	ISM02.4	PPB	11/22/2019 10:00 SMPL
K5913-01	MJLNT4	590	0.1111 -	ISM02.4	PPB	11/22/2019 10:02 SMPL
K5913-02	MJLNT5	951	0.2128 -	ISM02.4	PPB	11/22/2019 10:04 SMPL
K5913-03	MJLNT6	935	0.2083 -	ISM02.4	PPB	11/22/2019 10:07 SMPL
K5913-04	MJLNT7	802	0.1709 -	ISM02.4	PPB	11/22/2019 10:09 SMPL
K5913-05	MJLNT7D	806	0.172 -	ISM02.4	PPB	11/22/2019 10:11 SMPL
K5913-06	MJLNT7S	11473	3.1768 -	ISM02.4	PPB	11/22/2019 10:14 SMPL
K5913-07	MJLNT8	416	0.0621 -	ISM02.4	PPB	11/22/2019 10:16 SMPL
K5913-08	MJLNT9	1184	0.2785 -	ISM02.4	PPB	11/22/2019 10:18 SMPL
K5913-09	MJLNW0	853	0.1852 -	ISM02.4	PPB	11/22/2019 10:20 SMPL
K5913-10	MJLNW1	1194	0.2813 -	ISM02.4	PPB	11/22/2019 10:23 SMPL
K5913-11	MJLNW2	658	0.1303 -	ISM02.4	PPB	11/22/2019 10:25 SMPL
K5913-12	MJLNW3	788	0.1669 -	ISM02.4	PPB	11/22/2019 10:27 SMPL
K5913-13	MJLNW4	984	0.2221 -	ISM02.4	PPB	11/22/2019 10:29 SMPL
K5913-14	MJLNW5	504	0.0869 -	ISM02.4	PPB	11/22/2019 10:32 SMPL
K5913-15	MJLNW6	851	0.1847 -	ISM02.4	PPB	11/22/2019 10:34 SMPL
K5913-16	MJLNW7	499	0.0855 -	ISM02.4	PPB	11/22/2019 10:36 SMPL
K5913-17	MJLNW8	4466	1.203 -	ISM02.4	PPB	11/22/2019 10:39 SMPL
K5913-18	MJLNW9	1410	0.3421 -	ISM02.4	PPB	11/22/2019 10:41 SMPL
CCV013	CCV013	18652	5.1991 -	ISM02.4	PPB	11/22/2019 10:43 SMPL
CCB013	CCB013	68	-0.0359 -	ISM02.4	PPB	11/22/2019 10:45 SMPL
K5913-19	MJLNX0	978	0.2204 -	ISM02.4	PPB	11/22/2019 10:48 SMPL
CCV014	CCV014	18492	5.1541 -	ISM02.4	PPB	11/22/2019 10:50 SMPL
CCB014	CCB014	3	-0.0542 -	ISM02.4	PPB	11/22/2019 10:52 SMPL