

LB 106433


7471B

INSTRUMENT ID CV1

Linear

 μ Abs.:

42312



A= 0.0000e+000

B= 2.3333e-004

C= -1.5675e-002

Rho= 0.9995448

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.004	-0.004	48	0.000	48				
0.2	0.200	0.187	-0.013	870	0.0 %	870				
2.5	2.500	2.500	0.000	10783	0.0 %	10783				
5.0	5.000	4.937	-0.063	21224	0.0 %	21224				
7.5	7.500	7.723	0.223	33166	0.0 %	33166				
10.0	10.000	9.857	-0.143	42312	0.0 %	42312				

LB106433

Method: 7471B		Operator: Admin		Date of Analysis: 22 Nov 2019 11:22:34					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	48	-	0.0000	7471B	PPB	22 Nov 2019 11:24:07	S	Std
0.2 - 1	50.2	870	-	0.2000	7471B	PPB	22 Nov 2019 11:26:24	S	Std
2.5 - 1	52.5	10783	-	2.5000	7471B	PPB	22 Nov 2019 11:28:41	S	Std
5.0 - 1	55.0	21224	-	5.0000	7471B	PPB	22 Nov 2019 11:30:58	S	Std
7.5 - 1	57.5	33166	-	7.5000	7471B	PPB	22 Nov 2019 11:33:15	S	Std
10.0 - 1	510.0	42312	-	10.0000	7471B	PPB	22 Nov 2019 11:43:08	S	Std
ICV05 - 1	ICV05	18143	4.2177	-	7471B	PPB	22 Nov 2019 11:46:32	U	SMPL
ICB05 - 1	ICB05	-45	-0.0262	-	7471B	PPB	22 Nov 2019 11:48:50	U	SMPL
CCV19 - 1	CCV19	20369	4.7371	-	7471B	PPB	22 Nov 2019 11:51:06	U	SMPL
CCB19 - 1	CCB19	-34	-0.0236	-	7471B	PPB	22 Nov 2019 11:53:22	U	SMPL
CRA - 1	CRA	882	0.1901	-	7471B	PPB	22 Nov 2019 11:55:39	U	SMPL
HighStd - 1	HighStd	42039	9.7934	-	7471B	PPB	22 Nov 2019 11:57:54	U	SMPL
ChkStd - 1	ChkStd	29841	6.9472	-	7471B	PPB	22 Nov 2019 12:00:09	U	SMPL
PB125035BL - 1	PBS	-85	-0.0355	-	7471B	PPB	22 Nov 2019 12:02:27	U	SMPL
PB125035BS - 1	LCSS	18548	4.3122	-	7471B	PPB	22 Nov 2019 12:04:45	U	SMPL
K5670-09 - 1	OK-01-112119	2955	0.6738	-	7471B	PPB	22 Nov 2019 12:07:01	U	SMPL
K5671-11 - 1	OR-02-112119	3750	0.8593	-	7471B	PPB	22 Nov 2019 12:09:18	U	SMPL
K5676-21 - 1	SU-02-112019	3422	0.7828	-	7471B	PPB	22 Nov 2019 12:11:35	U	SMPL
K5676-21DUP - 1	SU-02-112019DUP	3404	0.7786	-	7471B	PPB	22 Nov 2019 12:13:53	U	SMPL
K5676-21MS - 1	SU-02-112019MS	18857	4.3843	-	7471B	PPB	22 Nov 2019 12:16:08	U	SMPL
CCV20 - 1	CCV20	22137	5.1496	-	7471B	PPB	22 Nov 2019 12:18:24	U	SMPL
CCB20 - 1	CCB20	-64	-0.0306	-	7471B	PPB	22 Nov 2019 12:20:41	U	SMPL
K5676-21MSD - 1	SU-02-112019MSD	18796	4.3701	-	7471B	PPB	22 Nov 2019 12:22:57	U	SMPL
K5676-26 - 1	SU-03-112019	3452	0.7898	-	7471B	PPB	22 Nov 2019 12:25:12	U	SMPL
K5948-01 - 1	B-1-0-6	3244	0.7413	-	7471B	PPB	22 Nov 2019 12:27:29	U	SMPL
K5948-02 - 1	B-1-6-12	398	0.0772	-	7471B	PPB	22 Nov 2019 12:29:44	U	SMPL
K5951-01 - 1	IRONBOUND-4FT	773	0.1647	-	7471B	PPB	22 Nov 2019 12:32:01	U	SMPL
K5951-03 - 1	IRONBOUND-8FT	381	0.0732	-	7471B	PPB	22 Nov 2019 12:34:17	U	SMPL
K5956-01 - 1	FAIRMOUNT-FOUNDATION-3	1013	0.2207	-	7471B	PPB	22 Nov 2019 12:36:34	U	SMPL
K5959-01 - 1	WS-112019-0-2-COMP-B3	18140	4.2170	-	7471B	PPB	22 Nov 2019 12:38:51	U	SMPL
K5959-03 - 1	WS-112019-2-5-COMP-B3	205	0.0322	-	7471B	PPB	22 Nov 2019 12:41:08	U	SMPL
K5959-06 - 1	WS-112019-0-2-COMP-B4	4794	1.1029	-	7471B	PPB	22 Nov 2019 12:43:24	U	SMPL
CCV21 - 1	CCV21	20988	4.8815	-	7471B	PPB	22 Nov 2019 12:45:40	U	SMPL
CCB21 - 1	CCB21	-44	-0.0259	-	7471B	PPB	22 Nov 2019 12:47:56	U	SMPL
K5959-07 - 1	WS-112019-2-5-COMP-B4	19249	4.4758	-	7471B	PPB	22 Nov 2019 12:50:13	U	SMPL
K5959-10 - 1	WS-112019-0-2-COMP-B2	93898	21.8939	-	7471B	PPB	22 Nov 2019 12:52:29	U	SMPL
K5959-11 - 1	WS-112019-2-5-COMP-B2	70756	16.4941	-	7471B	PPB	22 Nov 2019 12:54:45	U	SMPL
K5959-16 - 1	WS-112019-COMP-BD	171	0.0242	-	7471B	PPB	22 Nov 2019 12:57:04	U	SMPL
K5676-21LX5 - 1		751	0.1596	-	7471B	PPB	22 Nov 2019 12:59:23	U	SMPL
CCV22 - 1	CCV22	21701	5.0479	-	7471B	PPB	22 Nov 2019 13:01:39	U	SMPL
CCB22 - 1	CCB22	-104	-0.0399	-	7471B	PPB	22 Nov 2019 13:03:55	U	SMPL
K5959-10DLX3 - 1		32496	7.5667	-	7471B	PPB	22 Nov 2019 13:06:14	U	SMPL
K5959-11DLX3 - 1		24185	5.6275	-	7471B	PPB	22 Nov 2019 13:08:31	U	SMPL
CCV23 - 1	CCV23	19575	4.5518	-	7471B	PPB	22 Nov 2019 13:10:48	U	SMPL
CCB23 - 1	CCB23	-60	-0.0297	-	7471B	PPB	22 Nov 2019 13:13:06	U	SMPL