


LB106438
INSTRUMENT ID CV1

7470A



Linear

A= 0.0000e+000

B= 3.4086e-004

C= 2.8046e-002

Rho= 0.9996397

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.041	0.041	38	0.000	38				
0.2	0.200	0.259	0.059	678	0.0 %	678				
2.5	2.500	2.426	-0.074	7034	0.0 %	7034				
5.0	5.000	5.012	0.012	14621	0.0 %	14621				
7.5	7.500	7.327	-0.173	21415	0.0 %	21415				
10.0	10.000	10.135	0.135	29652	0.0 %	29652				

LB106438 INSTRUMENT ID CV1

Method: 7470A		Operator: Admin		Date of Analysis: 22 Nov 2019 13:54:19					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	38	-	0.0000	7470A	PPB	22 Nov 2019 13:55:56	S	Std
0.2 - 1	30.2	678	-	0.2000	7470A	PPB	22 Nov 2019 13:58:04	S	Std
2.5 - 1	52.5	7034	-	2.5000	7470A	PPB	22 Nov 2019 14:00:12	S	Std
5.0 - 1	65.0	14621	-	5.0000	7470A	PPB	22 Nov 2019 14:02:20	S	Std
7.5 - 1	57.5	21415	-	7.5000	7470A	PPB	22 Nov 2019 14:04:28	S	Std
10.0 - 1	510.0	29652	-	10.0000	7470A	PPB	22 Nov 2019 14:06:34	S	Std
ICV06 - 1	ICV06	12582	4.3167	-	7470A	PPB	22 Nov 2019 14:13:43	U	SMPL
ICB06 - 1	ICB06	-51	0.0107	-	7470A	PPB	22 Nov 2019 14:15:49	U	SMPL
CCV24 - 1	CCV24	15011	5.1446	-	7470A	PPB	22 Nov 2019 14:17:55	U	SMPL
CCB24 - 1	CCB24	-31	0.0175	-	7470A	PPB	22 Nov 2019 14:20:01	U	SMPL
CRA - 1	CRA	644	0.2476	-	7470A	PPB	22 Nov 2019 14:22:07	U	SMPL
HighStd - 1	HighStd	29718	10.1576	-	7470A	PPB	22 Nov 2019 14:24:13	U	SMPL
ChkStd - 1	ChkStd	20641	7.0636	-	7470A	PPB	22 Nov 2019 14:26:20	U	SMPL
PB125045BL - 1	PBW	-58	0.0083	-	7470A	PPB	22 Nov 2019 14:28:27	U	SMPL
K5668-11 - 1	HD-01-112119	-41	0.0141	-	7470A	PPB	22 Nov 2019 14:32:39	U	SMPL
K5668-11DUP - 1	HD-01-112119DUP	-31	0.0175	-	7470A	PPB	22 Nov 2019 14:34:47	U	SMPL
K5668-11MS - 1	HD-01-112119MS	13072	4.4837	-	7470A	PPB	22 Nov 2019 14:36:54	U	SMPL
K5668-11MSD - 1	HD-01-112119MSD	13074	4.4844	-	7470A	PPB	22 Nov 2019 14:39:03	U	SMPL
K5670-12 - 1	OK-01-112119	-35	0.0161	-	7470A	PPB	22 Nov 2019 14:41:09	U	SMPL
CCV25 - 1	CCV25	14688	5.0345	-	7470A	PPB	22 Nov 2019 14:43:15	U	SMPL
CCB25 - 1	CCB25	197	0.0952	-	7470A	PPB	22 Nov 2019 14:45:22	U	SMPL
K5671-15 - 1	OR-02-112119	10	0.0315	-	7470A	PPB	22 Nov 2019 14:47:28	U	SMPL
K5970-04 - 1	TP-28-S	33	0.0393	-	7470A	PPB	22 Nov 2019 14:49:34	U	SMPL
K5970-08 - 1	TP-28-D	32	0.0390	-	7470A	PPB	22 Nov 2019 14:51:40	U	SMPL
K5972-03 - 1	25441-26560	10	0.0315	-	7470A	PPB	22 Nov 2019 14:53:47	U	SMPL
K5972-05 - 1	SOIL-PILE	0	0.0280	-	7470A	PPB	22 Nov 2019 14:55:54	U	SMPL
K5973-03 - 1	DUCT-BANK-1	-7	0.0257	-	7470A	PPB	22 Nov 2019 14:58:01	U	SMPL
K5973-06 - 1	DUCT-BANK-2	-4	0.0267	-	7470A	PPB	22 Nov 2019 15:00:09	U	SMPL
K5973-09 - 1	TP-7	-16	0.0226	-	7470A	PPB	22 Nov 2019 15:02:16	U	SMPL
PB125046BL - 1	PBW	223	0.1041	-	7470A	PPB	22 Nov 2019 15:04:25	U	SMPL
PB125046BS - 1	LCSW	12622	4.3303	-	7470A	PPB	22 Nov 2019 15:06:31	U	SMPL
CCV26 - 1	CCV26	14865	5.0949	-	7470A	PPB	22 Nov 2019 15:08:37	U	SMPL
CCB26 - 1	CCB26	188	0.0921	-	7470A	PPB	22 Nov 2019 15:10:43	U	SMPL
K5676-25 - 1	SU-02-112019	12	0.0321	-	7470A	PPB	22 Nov 2019 15:12:50	U	SMPL
K5676-25DUP - 1	SU-02-112019DUP	75	0.0536	-	7470A	PPB	22 Nov 2019 15:14:56	U	SMPL
K5676-25MS - 1	SU-02-112019MS	12103	4.1534	-	7470A	PPB	22 Nov 2019 15:17:02	U	SMPL
K5676-25MSD - 1	SU-02-112019MSD	12409	4.2577	-	7470A	PPB	22 Nov 2019 15:19:08	U	SMPL
K5676-30 - 1	SU-03-112019	-30	0.0178	-	7470A	PPB	22 Nov 2019 15:21:15	U	SMPL
K5948-03 - 1	BSW-1-0-12	-6	0.0260	-	7470A	PPB	22 Nov 2019 15:23:23	U	SMPL
K5948-04 - 1	B-1-0-6	-12	0.0240	-	7470A	PPB	22 Nov 2019 15:25:29	U	SMPL
K5948-05 - 1	B-1-6-12	-16	0.0226	-	7470A	PPB	22 Nov 2019 15:27:37	U	SMPL
K5951-02 - 1	IRONBOUND-4FT	-12	0.0240	-	7470A	PPB	22 Nov 2019 15:29:45	U	SMPL
K5951-04 - 1	IRONBOUND-8FT	-5	0.0263	-	7470A	PPB	22 Nov 2019 15:31:51	U	SMPL
CCV27 - 1	CCV27	14162	4.8552	-	7470A	PPB	22 Nov 2019 15:33:57	U	SMPL
CCB27 - 1	CCB27	-24	0.0199	-	7470A	PPB	22 Nov 2019 15:36:03	U	SMPL
K5668-11LX5 - 1		7	0.0304	-	7470A	PPB	22 Nov 2019 15:38:10	U	SMPL
PB125046TB - 1	PB124988TB	26	0.0369	-	7470A	PPB	22 Nov 2019 15:40:16	U	SMPL
K5676-25LX5 - 1		41	0.0420	-	7470A	PPB	22 Nov 2019 15:42:22	U	SMPL
PB125045TB - 1	PB125014TB	22	0.0355	-	7470A	PPB	22 Nov 2019 15:44:28	U	SMPL
PB125045BS - 1	LCSW	12559	4.3088	-	7470A	PPB	22 Nov 2019 15:46:36	U	SMPL
CCV28 - 1	CCV28	14224	4.8764	-	7470A	PPB	22 Nov 2019 15:48:43	U	SMPL
CCB28 - 1	CCB28	-50	0.0110	-	7470A	PPB	22 Nov 2019 15:50:50	U	SMPL