



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106450

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/22/2019	17:40
2	CAL2	1	Water	NA	NA	20.2	7.00	11/22/2019	17:41
3	CAL3	1	Water	NA	NA	20.2	10.03	11/22/2019	17:42
4	ICV	1	Water	NA	NA	20.2	7.01	11/22/2019	17:44
5	CCV1	1	Water	NA	NA	20.2	2.01	11/22/2019	17:45
6	K5663-04	1	Solid	20.03	20	20.9	8.34	11/22/2019	17:48
7	K5672-05	1	Solid	20.04	20	21.4	7.69	11/22/2019	17:50
8	K5676-34	1	Solid	20.02	20	20.6	8.72	11/22/2019	18:00
9	K6001-03	1	Solid	20.03	20	20.7	8.73	11/22/2019	18:01
10	K6001-03DUP	1	Solid	20.04	20	20.8	8.74	11/22/2019	18:02
11	CCV2	1	Water	NA	NA	20.1	12.02	11/22/2019	18:05

WORKLIST(Hardcopy Internal Chain)

VB 106450

WorkList Name : CORROSIONITY K5678

WorkList ID : 133625

Department : Wet-Chemistry

Date : 11-22-2019 16:15:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/22/2019	Solid	K5663-04	Corrosivity	Cool 4 deg C	PSEG05	N12	AU-06-112219	11/22/2019	9045D
11/22/2019	Solid	K5672-05	Corrosivity	Cool 4 deg C	PSEG05	N12	EO-02-112219	11/22/2019	9045D
11/22/2019	Solid	K5676-34	Corrosivity	Cool 4 deg C	PSEG05		SU-04-112219	11/22/2019	9045D
11/21/2019	Solid	K6001-03	Corrosivity	Cool 4 deg C	PSEG01	I11	TP-12B-D	11/21/2019	9045D

Date/Time 11-22-19 16:15
 Received by: J.
 Relinquished by: ea

Date/Time 11-22-19 18:00
 Received by: CF
 Relinquished by: Jp