



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106461

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/25/2019	10:50
2	CAL2	1	Water	NA	NA	20.1	6.99	11/25/2019	10:51
3	CAL3	1	Water	NA	NA	20.1	10.03	11/25/2019	10:53
4	ICV	1	Water	NA	NA	20.1	7.00	11/25/2019	11:00
5	CCV1	1	Water	NA	NA	20.3	2.02	11/25/2019	11:01
6	K6011-01	1	Water	NA	NA	15.1	4.43	11/25/2019	11:03
7	K6011-01DUP	1	Water	NA	NA	15.2	4.44	11/25/2019	11:04
8	CCV2	1	Water	NA	NA	20.1	12.02	11/25/2019	11:05

LB 106461

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K6011 WorkList ID : 133666 Department : Wet-Chemistry Date : 11-25-2019 10:41:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/21/2019	Water	K6011-01 - 1	pH	Cool 4 deg C	STER14	I51	MINI-CELL	11/21/2019	SM 4500-H B

Date/Time 11-25-19 10:45
Received by: JD
Relinquished by: JD

Date/Time 11-25-19 11:30
Received by: CD
Relinquished by: JD