



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 11/26/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 18:00
In Date: 11/25/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 11/26/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106469

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
K5667-01	NB-07-112519	1	1.14	9.51	8.6	89.1	
K5667-02	NB-07-112519-1	2	1.14	9.66	8.62	87.8	
K5667-03	NB-07-112519-2	3	1.16	9.91	8.73	86.5	
K5908-15	B-1-(12-18)	4	1.15	9.89	6.96	66.5	
K5908-17	B-8-2-(12-18)	5	1.13	9.78	8.69	87.4	
K6001-03	TP-12B-D	6	1.17	9.78	9.08	91.9	
K6006-01	01-A-01-B-01-C	7	1.00	2.00	2.00	100.0	CLUCK SAMPLE 100% SOLIDS
K6006-02	02-A-02-B-02-C	8	1.00	2.00	2.00	100.0	CLUCK SAMPLE 100% SOLIDS
K6006-03	03-A-03-B-03-C	9	1.00	2.00	2.00	100.0	CLUCK SAMPLE 100% SOLIDS
K6006-04	04-A-04-B-04-C	10	1.00	2.00	2.00	100.0	CLUCK SAMPLE 100% SOLIDS
K6006-05	05-A-05-B-05-C	11	1.00	2.00	2.00	100.0	CLUCK SAMPLE 100% SOLIDS
K6010-01	CALVSOIL1-20191121	12	1.14	9.6	9.39	97.5	
K6012-01	211	13	1.15	9.98	8.51	83.4	
K6013-01	600436752	14	1.13	9.71	8.42	85.0	
K6013-02	600436752	15	1.15	9.7	8.21	82.6	
K6014-01	CS-01	16	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
K6014-03	CS-02	17	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
K6014-05	CS-03	18	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
K6014-07	CS-04	19	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
K6014-09	CS-05	20	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
K6014-11	CS-06	21	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
K6014-13	CS-07	22	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
K6014-15	CS-08	23	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
K6014-17	CS-09	24	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
K6018-01	ETGI-359	25	1.14	9.68	7.66	76.3	
K6019-01	BUS-SUPPORT-3/5	26	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
K6021-01	SOIL-001-E	27	1.12	9.64	8.88	91.1	
K6021-04	SOIL-002-E	28	1.12	9.7	9.24	94.6	



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 11/26/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 18:00
In Date: 11/25/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 11/26/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106469

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K6021-05	SOIL-003-E	29	1.15	9.75	8.83	89.3	
K6021-06	SOIL-004-E	30	1.14	9.8	8.76	88.0	
K6022-01	TP-15-S1	31	1.15	9.68	9.22	94.6	
K6022-02	TP-15-S2	32	1.13	9.8	9.34	94.7	
K6022-03	TP-15-S	33	1.14	9.56	9.14	95.0	
K6022-05	TP-16-S1	34	1.14	9.53	8.77	90.9	
K6022-06	TP-16-S2	35	1.15	9.57	8.8	90.9	
K6022-07	TP-16-S	36	1.16	9.85	8.94	89.5	
K6022-09	TP-16-D1	37	1.14	9.77	8.47	84.9	
K6022-10	TP-16-D2	38	1.16	9.7	9.36	96.0	
K6022-11	TP-16-D	39	1.15	9.72	9.37	95.9	
K6022-13	TP-17-S1	40	1.13	9.9	9.32	93.4	
K6022-14	TP-17-S2	41	1.12	9.93	9.64	96.7	
K6022-15	TP-17-S	42	1.12	9.56	9.36	97.6	
K6024-01	B-1-(0.5-1.0)	43	1.13	9.75	8.37	84.0	
K6024-02	B-2-(0.5-1.0)	44	1.12	9.67	7.91	79.4	
K6024-03	B-3-(0.5-1.0)	45	1.15	9.7	8.15	81.9	
K6024-04	B-4-(0.5-1.0)	46	1.14	9.78	8.33	83.2	
K6024-05	B-5-(0.5-1.0)	47	1.12	9.71	8.26	83.1	
K6024-06	B-6-(0.5-1.0)	48	1.13	9.94	8.63	85.1	
K6025-01	SOIL-CUTTINGS	49	1.15	9.76	9.37	95.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$