



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106470

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/25/2019	18:00
2	CAL2	1	Water	NA	NA	20.3	7.01	11/25/2019	18:01
3	CAL3	1	Water	NA	NA	20.3	10.03	11/25/2019	18:02
4	ICV	1	Water	NA	NA	20.2	6.99	11/25/2019	18:05
5	CCV1	1	Water	NA	NA	20.2	2.02	11/25/2019	18:06
6	K5667-04	1	Solid	20.03	20	21.5	7.44	11/25/2019	18:07
7	K6001-04	1	Solid	20.04	20	20.7	8.73	11/25/2019	18:08
8	K6010-01	1	Solid	20.04	20	22.4	5.72	11/25/2019	18:10
9	K6013-03	1	Solid	20.02	20	21.3	8.87	11/25/2019	18:11
10	K6022-04	1	Solid	20.04	20	22.0	8.41	11/25/2019	18:13
11	K6022-08	1	Solid	20.02	20	21.9	8.02	11/25/2019	18:15
12	K6022-12	1	Solid	20.03	20	21.9	7.30	11/25/2019	18:16
13	K6022-16	1	Solid	20.04	20	21.5	9.04	11/25/2019	18:20
14	K6025-01	1	Solid	20.03	20	21.8	5.61	11/25/2019	18:22
15	K6025-01DUP	1	Solid	20.04	20	21.9	5.62	11/25/2019	18:23
16	CCV2	1	Water	NA	NA	20.2	12.01	11/25/2019	18:25

106470

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K6010 WorkList ID : 133678 Department : Wet-Chemistry Date : 11-25-2019 16:03:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/25/2019	Solid	K5667-04	Corrosivity	Cool 4 deg C	PSEG05	N12	NB-07-112519	11/25/2019	9045D
11/21/2019	Solid	K6001-04	Corrosivity	Cool 4 deg C	PSEG01		TP-12B-D	11/21/2019	9045D
11/21/2019	Solid	K6010-01	Corrosivity	Cool 4 deg C	TETR06	I51	CALVSOIL 1-20191121	11/21/2019	9045D
11/25/2019	Solid	K6013-03	Corrosivity	Cool 4 deg C	PSEG01	N12	600436752	11/25/2019	9045D
11/25/2019	Solid	K6022-04	Corrosivity	Cool 4 deg C	PSEG01	N12	TP-15-S	11/25/2019	9045D
11/25/2019	Solid	K6022-08	Corrosivity	Cool 4 deg C	PSEG01	N12	TP-16-S	11/25/2019	9045D
11/25/2019	Solid	K6022-12	Corrosivity	Cool 4 deg C	PSEG01	N12	TP-16-D	11/25/2019	9045D
11/25/2019	Solid	K6022-16	Corrosivity	Cool 4 deg C	PSEG01	N12	TP-17-S	11/25/2019	9045D
11/25/2019	Solid	K6025-01	Corrosivity	Cool 4 deg C	SEAC01	I41	SOIL-CUTTINGS	11/25/2019	9045D

Date/Time 11-25-19 16:10
 Received by: Ja
 Relinquished by: cd

Date/Time 11-25-19 18:00
 Received by: cd
 Relinquished by: Jo