



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106471

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/25/2019	18:00
2	CAL2	1	Water	NA	NA	20.3	7.01	11/25/2019	18:01
3	CAL3	1	Water	NA	NA	20.3	10.03	11/25/2019	18:02
4	ICV	1	Water	NA	NA	20.3	6.99	11/25/2019	18:05
5	CCV1	1	Water	NA	NA	20.3	2.01	11/25/2019	18:30
6	K6025-02	1	Water	NA	NA	19.2	7.11	11/25/2019	18:31
7	K6025-02DUP	1	Water	NA	NA	19.1	7.12	11/25/2019	18:33
8	CCV2	1	Water	NA	NA	20.2	12.01	11/25/2019	18:35

28106471

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY K6025 WorkList ID : 133681 Department : Wet-Chemistry Date : 11-25-2019 16:55:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/25/2019	Water	K6025-02	Corrosivity	Cool 4 deg C	SEAC01	I41	GROUNDWATER	11/25/2019	9040C

Date/Time 11.25.19 1610
 Received by: JD
 Relinquished by: CD

Date/Time 11.25.19 18.00
 Received by: CD
 Relinquished by: JD