



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: jignesh

Date: 11/26/2019

Run Number: LB106477

BalanceID: WC SC-3

OvenID: WC-OVEN-2

FilterID: 1827-047

ThermometerID: WET OVEN#1

TEMP1 IN: 103 °C 11/26/2019 07:10 TEMP1 OUT: 103 °C 11/26/2019 08:10
TEMP2 IN: 103 °C 11/26/2019 08:45 TEMP2 OUT: 103 °C 11/26/2019 10:00
TEMP3 IN: 103 °C 11/26/2019 11:15 TEMP3 OUT: 103 °C 11/26/2019 12:40
TEMP4 IN: 103 °C 11/26/2019 13:20 TEMP4 OUT: 103 °C 11/26/2019 15:35

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB106477BL	LB106477BL	1.3487	1.3487	100	1.3487	1.3487	1.3487	0.0000	0
2	LB106477BS	LB106477BS	1.4036	1.4036	100	1.4581	1.4581	1.4581	0.0545	545
3	K6015-01	112519-JETT-F-D-B	1.3845	1.3845	300	1.4176	1.4176	1.4176	0.0331	110.3
4	K6015-02	112519-JETT-F-D-M	1.3823	1.3823	300	1.4099	1.4099	1.4099	0.0276	92
5	K6015-03	112519-JETT-F-D-S	1.3778	1.3778	300	1.4085	1.4085	1.4085	0.0307	102.3
6	K6015-04	112519-JETT-F-U-B	1.3891	1.3891	300	1.4228	1.4228	1.4228	0.0337	112.3
7	K6015-05	112519-JETT-F-U-M	1.3791	1.3791	300	1.4011	1.4011	1.4011	0.0220	73.3
8	K6015-06	112519-JETT-F-U-S	1.3911	1.3911	300	1.4053	1.4053	1.4053	0.0142	47.3
9	K6015-07	112519-DBRI-F-U-B	1.3931	1.3931	300	1.4251	1.4251	1.4251	0.0320	106.7
10	K6015-08	112519-DBRI-F-U-M	1.3841	1.3841	300	1.4077	1.4077	1.4077	0.0236	78.7
11	K6015-09	112519-DBRI-F-U-S	1.3789	1.3789	300	1.3935	1.3935	1.3935	0.0146	48.7
12	K6015-10	112519-DBRI-F-D-B	1.3758	1.3758	300	1.4057	1.4057	1.4057	0.0299	99.7
13	K6015-11	112519-DBRI-F-D-M	1.3909	1.3909	300	1.4191	1.4191	1.4191	0.0282	94
14	K6015-12	112519-DBRI-F-D-S	1.3913	1.3913	300	1.4235	1.4235	1.4235	0.0322	107.3
15	K6015-12DUP	112519-DBRI-F-D-SDUP	1.3841	1.3841	300	1.4166	1.4166	1.4166	0.0325	108.3

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: jignesh

Date: 11/26/2019

Run Number: LB106477

BalanceID: WC SC-3

OvenID: WC-OVEN-2

FilterID: 1827-047

ThermometerID: WET OVEN#1

TEMP1 IN: 103 °C 11/26/2019 07:10 TEMP1 OUT: 103 °C 11/26/2019 08:10
 TEMP2 IN: 103 °C 11/26/2019 08:45 TEMP2 OUT: 103 °C 11/26/2019 10:00
 TEMP3 IN: 103 °C 11/26/2019 11:15 TEMP3 OUT: 103 °C 11/26/2019 12:40
 TEMP4 IN: 103 °C 11/26/2019 13:20 TEMP4 OUT: 103 °C 11/26/2019 15:35

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	K6016-03	HP-627	1.3997	1.3997	300	1.4150	1.4150	1.4150	0.0153	51
17	K6017-01	112219-DBRI-E-D-B	1.3795	1.3795	800	1.4045	1.4045	1.4045	0.0250	31.3
18	K6017-02	112219-DBRI-E-D-M	1.3830	1.3830	800	1.4012	1.4012	1.4012	0.0182	22.8
19	K6017-03	112219-DBRI-E-D-S	1.3953	1.3953	800	1.4044	1.4044	1.4044	0.0091	11.4
20	K6017-04	112219-DBRI-E-U-B	1.3953	1.3953	800	1.4193	1.4193	1.4193	0.0240	30
21	K6017-05	112219-DBRI-E-U-M	1.4033	1.4033	800	1.4217	1.4217	1.4217	0.0184	23
22	K6017-06	112219-DBRI-E-U-S	1.3932	1.3932	800	1.4059	1.4059	1.4059	0.0127	15.9
23	K6017-07	112219-TIPC-E-D-M	1.3966	1.3966	800	1.4240	1.4240	1.4240	0.0274	34.3
24	K6017-08	112219-TIPC-E-U-M	1.3757	1.3757	500	1.4068	1.4068	1.4068	0.0311	62.2
25	K6017-08DUP	112219-TIPC-E-U-MDUP	1.3793	1.3793	500	1.4093	1.4093	1.4093	0.0300	60

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

28106477

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS K6015 WorkList ID : 133684 Department : Wet-Chemistry Date : 11-26-2019 07:59:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/02/2019	Water	K6015-01 ✓	TSS	Cool 4 deg C	HDR103	I51	112519-JETT-F-D-B	11/25/2019	SM2540 D
12/02/2019	Water	K6015-02 ✓	TSS	Cool 4 deg C	HDR103	I51	112519-JETT-F-D-M	11/25/2019	SM2540 D
12/02/2019	Water	K6015-03 ✓	TSS	Cool 4 deg C	HDR103	I51	112519-JETT-F-D-S	11/25/2019	SM2540 D
12/02/2019	Water	K6015-04 ✓	TSS	Cool 4 deg C	HDR103	I51	112519-JETT-F-U-B	11/25/2019	SM2540 D
12/02/2019	Water	K6015-05 ✓	TSS	Cool 4 deg C	HDR103	I51	112519-JETT-F-U-M	11/25/2019	SM2540 D
12/02/2019	Water	K6015-06 ✓	TSS	Cool 4 deg C	HDR103	I51	112519-JETT-F-U-S	11/25/2019	SM2540 D
12/02/2019	Water	K6015-07 ✓	TSS	Cool 4 deg C	HDR103	I51	112519-DBRI-F-U-B	11/25/2019	SM2540 D
12/02/2019	Water	K6015-08 ✓	TSS	Cool 4 deg C	HDR103	I51	112519-DBRI-F-U-M	11/25/2019	SM2540 D
12/02/2019	Water	K6015-09 ✓	TSS	Cool 4 deg C	HDR103	I51	112519-DBRI-F-U-S	11/25/2019	SM2540 D
12/02/2019	Water	K6015-10 ✓	TSS	Cool 4 deg C	HDR103	I51	112519-DBRI-F-D-B	11/25/2019	SM2540 D
12/02/2019	Water	K6015-11 ✓	TSS	Cool 4 deg C	HDR103	I51	112519-DBRI-F-D-M	11/25/2019	SM2540 D
12/02/2019	Water	K6015-12 ✓	TSS	Cool 4 deg C	HDR103	I51	112519-DBRI-F-D-S	11/25/2019	SM2540 D
12/01/2019	Water	K6016-03 -E	TSS	Cool 4 deg C	CAMP02		HP-627	11/24/2019	SM2540 D
11/29/2019	Water	K6017-01 ✓	TSS	Cool 4 deg C	HDR103	I51	112219-DBRI-E-D-B	11/22/2019	SM2540 D
11/29/2019	Water	K6017-02 ✓	TSS	Cool 4 deg C	HDR103	I51	112219-DBRI-E-D-M	11/22/2019	SM2540 D
11/29/2019	Water	K6017-03 ✓	TSS	Cool 4 deg C	HDR103	I51	112219-DBRI-E-D-S	11/22/2019	SM2540 D
11/29/2019	Water	K6017-04 ✓	TSS	Cool 4 deg C	HDR103	I51	112219-DBRI-E-U-B	11/22/2019	SM2540 D
11/29/2019	Water	K6017-05 ✓	TSS	Cool 4 deg C	HDR103	I51	112219-DBRI-E-U-M	11/22/2019	SM2540 D
11/29/2019	Water	K6017-06 ✓	TSS	Cool 4 deg C	HDR103	I51	112219-DBRI-E-U-S	11/22/2019	SM2540 D
11/29/2019	Water	K6017-07 ✓	TSS	Cool 4 deg C	HDR103	I51	112219-TIPC-E-D-M	11/22/2019	SM2540 D
11/29/2019	Water	K6017-08 ✓	TSS	Cool 4 deg C	HDR103	I51	112219-TIPC-E-U-M	11/22/2019	SM2540 D

Date/Time 11.26.19 10:00
 Received by: J2
 Relinquished by: [Signature]

Date/Time 11.26.19 11:45
 Received by: [Signature]
 Relinquished by: J2