


INSTRUMENT ID CV1
LB106487

7470A



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	19	0.000	19				
0.2	0.200	0.198	-0.002	806	0.0 %	806				
2.5	2.500	2.542	0.042	10530	0.0 %	10530				
5.0	5.000	4.917	-0.083	20385	0.0 %	20385				
7.5	7.500	7.518	0.018	31178	0.0 %	31178				
10.0	10.000	10.016	0.016	41543	0.0 %	41543				

LB106487 INSTRUMENT ID CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Nov 2019 13:08:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	19	-	0.0000	7470A	PPB	26 Nov 2019 13:17:09	S	Std
0.2 - 1	50.2	806	-	0.2000	7470A	PPB	26 Nov 2019 13:19:16	S	Std
2.5 - 1	52.5	10530	-	2.5000	7470A	PPB	26 Nov 2019 13:21:24	S	Std
5.0 - 1	56.0	20385	-	5.0000	7470A	PPB	26 Nov 2019 13:23:32	S	Std
7.5 - 1	57.5	31178	-	7.5000	7470A	PPB	26 Nov 2019 13:25:39	S	Std
10.0 - 1	510.0	41543	-	10.0000	7470A	PPB	26 Nov 2019 13:27:46	S	Std
ICV08 - 1	ICV08	16409	3.9587	-	7470A	PPB	26 Nov 2019 13:30:33	U	SMPL
ICB08 - 1	ICB08	-20	-0.0010	-	7470A	PPB	26 Nov 2019 13:32:39	U	SMPL
CCV36 - 1	CCV36	20522	4.9500	-	7470A	PPB	26 Nov 2019 13:34:45	U	SMPL
CCB36 - 1	CCB36	-50	-0.0082	-	7470A	PPB	26 Nov 2019 13:36:51	U	SMPL
CRA - 1	CRA	775	0.1906	-	7470A	PPB	26 Nov 2019 13:38:58	U	SMPL
HighStd - 1	HighStd	41329	9.9649	-	7470A	PPB	26 Nov 2019 13:41:04	U	SMPL
ChkStd - 1	ChkStd	27697	6.6793	-	7470A	PPB	26 Nov 2019 13:43:10	U	SMPL
PB125093BL - 1	PBW	-53	-0.0089	-	7470A	PPB	26 Nov 2019 13:45:18	U	SMPL
PB125093BS - 1	LCSW	16518	3.9850	-	7470A	PPB	26 Nov 2019 13:47:25	U	SMPL
K5959-13 - 1	FIELD-BLANK	3	0.0046	-	7470A	PPB	26 Nov 2019 13:49:32	U	SMPL
K5959-13DUP - 1	FIELD-BLANKDUP	31	0.0113	-	7470A	PPB	26 Nov 2019 13:51:39	U	SMPL
K5959-13MS - 1	FIELD-BLANKMS	16967	4.0932	-	7470A	PPB	26 Nov 2019 13:53:47	U	SMPL
K5959-13MSD - 1	FIELD-BLANKMSD	17063	4.1163	-	7470A	PPB	26 Nov 2019 13:55:55	U	SMPL
K5968-23 - 1	FB-03-20191121	-65	-0.0118	-	7470A	PPB	26 Nov 2019 13:58:01	U	SMPL
CCV37 - 1	CCV37	19989	4.8215	-	7470A	PPB	26 Nov 2019 14:00:08	U	SMPL
CCB37 - 1	CCB37	-36	-0.0048	-	7470A	PPB	26 Nov 2019 14:02:14	U	SMPL
K6008-01 - 1	CALVTANK-1-20191121	2	0.0043	-	7470A	PPB	26 Nov 2019 14:04:20	U	SMPL
PB125094BL - 1	PBW	25	0.0099	-	7470A	PPB	26 Nov 2019 14:06:26	U	SMPL
PB125094BS - 1	LCSW	16388	3.9536	-	7470A	PPB	26 Nov 2019 14:08:33	U	SMPL
K5663-04 - 1	AU-06-112219	0	0.0038	-	7470A	PPB	26 Nov 2019 14:10:40	U	SMPL
K5663-04DUP - 1	AU-06-112219DUP	1	0.0041	-	7470A	PPB	26 Nov 2019 14:12:47	U	SMPL
K5663-04MS - 1	AU-06-112219MS	18200	4.3904	-	7470A	PPB	26 Nov 2019 14:14:53	U	SMPL
K5663-04MSD - 1	AU-06-112219MSD	18232	4.3981	-	7470A	PPB	26 Nov 2019 14:17:01	U	SMPL
K5672-05 - 1	EO-02-112219	-65	-0.0118	-	7470A	PPB	26 Nov 2019 14:19:09	U	SMPL
K5676-34 - 1	SU-04-112219	7	0.0055	-	7470A	PPB	26 Nov 2019 14:21:17	U	SMPL
K5982-02 - 1	OILY-SPILL-PADS	32	0.0115	-	7470A	PPB	26 Nov 2019 14:23:23	U	SMPL
CCV38 - 1	CCV38	20432	4.9283	-	7470A	PPB	26 Nov 2019 14:25:29	U	SMPL
CCB38 - 1	CCB38	-18	-0.0005	-	7470A	PPB	26 Nov 2019 14:27:35	U	SMPL
K6001-04 - 1	TP-12B-D	16	0.0077	-	7470A	PPB	26 Nov 2019 14:29:42	U	SMPL
K5959-13LX5 - 1		8	0.0058	-	7470A	PPB	26 Nov 2019 14:31:49	U	SMPL
K5663-04LX5 - 1		-2	0.0034	-	7470A	PPB	26 Nov 2019 14:33:55	U	SMPL
PB125094TB - 1	PB125049TB	44	0.0144	-	7470A	PPB	26 Nov 2019 14:36:02	U	SMPL
CCV39 - 1	CCV39	20297	4.8958	-	7470A	PPB	26 Nov 2019 14:38:09	U	SMPL
CCB39 - 1	CCB39	-18	-0.0005	-	7470A	PPB	26 Nov 2019 14:40:16	U	SMPL