



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 11/27/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 17:34
In Date: 11/26/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 11/27/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106504

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
K5675-09	TR-05-112619	1	1.14	9.84	8.99	90.2	
K5675-10	TR-05-112619	2	1.15	9.93	9.11	90.7	
K6030-01	391119711	3	1.14	9.97	2.06	10.4	SLUDGE SAMPLE
K6053-01	EXT-084-SW178-112519	4	1.13	9.71	8.93	90.9	
K6053-02	EXT-084-SW178-112519-1	5	1.15	9.55	8.42	86.5	
K6053-03	EXT-084-SW179-112519	6	1.14	9.65	8.9	91.2	
K6053-04	EXT-084-SW179-112519-1	7	1.13	9.72	8.83	89.6	
K6053-05	EXT-029-SW180-112519	8	1.15	9.8	8.84	88.9	
K6053-06	EXT-029-SW181-112519	9	1.14	9.97	8.78	86.5	
K6054-01	RBR-50	10	1.12	9.96	8.63	85.0	
K6054-02	RBR-50	11	1.13	9.85	9.21	92.7	
K6054-04	RBR-28	12	1.13	9.64	8.86	90.8	
K6054-05	RBR-28	13	1.12	9.82	9.16	92.4	
K6062-01	HA-10	14	1.00	2.00	2.00	100.0	CLUCK SAMPLE 100% SOLIDS
K6062-02	HA-12	15	1.00	2.00	2.00	100.0	CLUCK SAMPLE 100% SOLIDS
K6062-03	HA-13	16	1.00	2.00	2.00	100.0	CLUCK SAMPLE 100% SOLIDS
K6062-04	HA-17	17	1.00	2.00	2.00	100.0	CLUCK SAMPLE 100% SOLIDS
K6062-05	HA-21	18	1.00	2.00	2.00	100.0	CLUCK SAMPLE 100% SOLIDS
K6065-01	RI-1	19	1.12	9.71	9.09	92.8	
K6065-02	RI-2	20	1.12	9.6	9.05	93.5	
K6065-03	RI-3	21	1.12	9.62	9.03	93.1	
K6065-04	RI-4	22	1.12	9.84	9.26	93.3	
K6065-05	RI-5	23	1.17	9.69	9.12	93.3	
K6066-01	TP-19-S1	24	1.14	9.81	9.12	92.0	
K6066-02	TP-19-S2	25	1.12	9.86	9.14	91.8	
K6066-03	TP-19-S	26	1.12	9.98	9.41	93.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$