



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106505

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/26/2019	17:40
2	CAL2	1	Water	NA	NA	20.1	7.01	11/26/2019	17:41
3	CAL3	1	Water	NA	NA	20.1	10.03	11/26/2019	17:43
4	ICV	1	Water	NA	NA	20.1	6.99	11/26/2019	17:45
5	CCV1	1	Water	NA	NA	20.2	2.01	11/26/2019	17:46
6	K5675-11	1	Solid	20.03	20	21.5	8.17	11/26/2019	17:47
7	K5675-11DUP	1	Solid	20.04	20	21.6	8.18	11/26/2019	17:48
8	K6054-03	1	Solid	20.02	20	22.9	7.89	11/26/2019	17:49
9	K6054-06	1	Solid	20.03	20	22.6	6.80	11/26/2019	17:50
10	K6066-04	1	Solid	20.04	20	21.8	9.31	11/26/2019	18:00
11	CCV2	1	Water	NA	NA	20.1	12.02	11/26/2019	18:02

106505

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K5675 WorkList ID : 133731 Department : Wet-Chemistry Date : 11-26-2019 15:06:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/26/2019	Solid	K5675-11	Corrosivity	Cool 4 deg C	PSEG05	N12	TR-05-112619	11/26/2019	9045D
11/26/2019	Solid	K6054-03	Corrosivity	Cool 4 deg C	PSEG01	N12	RBR-50	11/26/2019	9045D
11/26/2019	Solid	K6054-06	Corrosivity	Cool 4 deg C	PSEG01	N12	RBR-28	11/26/2019	9045D
11/26/2019	Solid	K6066-04	Corrosivity	Cool 4 deg C	PSEG01	N12	TP-19-S	11/26/2019	9045D

Date/Time 11.26.19 15:10
Received by: h
Relinquished by: h

Date/Time 11.26.19 18:00
Received by: h
Relinquished by: h