


LB106508
INSTRUMENT ID CV1

7471B



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.011	0.011	1	0.000	1				
0.2	0.200	0.264	0.064	1328	0.0 %	1328				
2.5	2.500	2.513	0.013	13133	0.0 %	13133				
5.0	5.000	5.232	0.232	27403	0.0 %	27403				
7.5	7.500	6.902	-0.598	36171	0.0 %	36171				
10.0	10.000	10.279	0.279	53892	0.0 %	53892				

LB106508 INSTRUMENT ID CV1

Method: 7471B		Operator: Admin		Date of Analysis: 27 Nov 2019 07:25:31					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	80.0	1	-	0.0000	7471B	PPB	27 Nov 2019 07:32:19	S	Std
0.2 - 1	50.2	1328	-	0.2000	7471B	PPB	27 Nov 2019 07:34:36	S	Std
2.5 - 1	32.5	13133	-	2.5000	7471B	PPB	27 Nov 2019 07:36:53	S	Std
5.0 - 1	56.6	27403	-	5.0000	7471B	PPB	27 Nov 2019 07:39:10	S	Std
7.5 - 1	57.5	36171	-	7.5000	7471B	PPB	27 Nov 2019 07:41:27	S	Std
10.0 - 1	510.0	53892	-	10.0000	7471B	PPB	27 Nov 2019 07:43:43	S	Std
ICV10 - 1	ICV10	22732	4.3418	-	7471B	PPB	27 Nov 2019 07:48:27	U	SMPL
ICB10 - 1	ICB10	-12	0.0084	-	7471B	PPB	27 Nov 2019 07:50:42	U	SMPL
CCV43 - 1	CCV43	27015	5.1578	-	7471B	PPB	27 Nov 2019 07:52:58	U	SMPL
CCB43 - 1	CCB43	-10	0.0087	-	7471B	PPB	27 Nov 2019 07:55:13	U	SMPL
CRA - 1	CRA	1268	0.2522	-	7471B	PPB	27 Nov 2019 08:00:42	U	SMPL
HighStd - 1	HighStd	53466	10.1975	-	7471B	PPB	27 Nov 2019 08:02:58	U	SMPL
ChkStd - 1	ChkStd	32937	6.2861	-	7471B	PPB	27 Nov 2019 08:05:13	U	SMPL
PB125145BL - 1	PBS	-32	0.0045	-	7471B	PPB	27 Nov 2019 08:07:30	U	SMPL
PB125145BS - 1	LCSS	22538	4.3048	-	7471B	PPB	27 Nov 2019 08:09:47	U	SMPL
K5675-09 - 1	TR-05-112619	844	0.1714	-	7471B	PPB	27 Nov 2019 08:12:04	U	SMPL
K6012-01 - 1	211	1484	0.2934	-	7471B	PPB	27 Nov 2019 08:14:21	U	SMPL
K6013-01 - 1	600436752	5441	1.0473	-	7471B	PPB	27 Nov 2019 08:16:37	U	SMPL
K6018-01 - 1	ETGI-359	27606	5.2704	-	7471B	PPB	27 Nov 2019 08:18:54	U	SMPL
K6022-03 - 1	TP-15-S	310	0.0697	-	7471B	PPB	27 Nov 2019 08:21:10	U	SMPL
CCV44 - 1	CCV44	26818	5.1203	-	7471B	PPB	27 Nov 2019 08:23:25	U	SMPL
CCB44 - 1	CCB44	15	0.0135	-	7471B	PPB	27 Nov 2019 08:25:41	U	SMPL
K6022-03DUP - 1	TP-15-SDUP	366	0.0804	-	7471B	PPB	27 Nov 2019 08:27:56	U	SMPL
K6022-03MS - 1	TP-15-SMS	20436	3.9043	-	7471B	PPB	27 Nov 2019 08:30:12	U	SMPL
K6022-03MSD - 1	TP-15-SMSD	20313	3.8809	-	7471B	PPB	27 Nov 2019 08:32:27	U	SMPL
K6022-07 - 1	TP-16-S	23263	4.4429	-	7471B	PPB	27 Nov 2019 08:34:44	U	SMPL
K6022-11 - 1	TP-16-D	433	0.0931	-	7471B	PPB	27 Nov 2019 08:37:00	U	SMPL
K6022-15 - 1	TP-17-S	1333	0.2646	-	7471B	PPB	27 Nov 2019 08:39:17	U	SMPL
K6024-01 - 1	B-1-(0.5-1.0)	100194	19.1005	-	7471B	PPB	27 Nov 2019 08:41:34	U	SMPL
K6024-02 - 1	B-2-(0.5-1.0)	358227	HIGH	-	7471B	PPB	27 Nov 2019 08:43:50	X	
K6024-03 - 1	B-3-(0.5-1.0)	394946	75.2595	-	7471B	PPB	27 Nov 2019 08:45:45	U	SMPL
K6024-04 - 1	B-4-(0.5-1.0)	1030696	HIGH	-	7471B	PPB	27 Nov 2019 08:53:09	X	
CCV45 - 1	CCV45	26145	4.9920	-	7471B	PPB	27 Nov 2019 08:54:56	U	SMPL
CCB45 - 1	CCB45	-210	-0.0294	-	7471B	PPB	27 Nov 2019 08:57:39	U	SMPL
K6024-05 - 1	B-5-(0.5-1.0)	345660	HIGH	-	7471B	PPB	27 Nov 2019 08:59:56	X	
K6024-06 - 1	B-6-(0.5-1.0)	106384	20.2799	-	7471B	PPB	27 Nov 2019 09:01:46	U	SMPL
K6054-01 - 1	RBR-50	21324	4.0735	-	7471B	PPB	27 Nov 2019 09:04:04	U	SMPL
K6054-04 - 1	RBR-28	1940	0.3803	-	7471B	PPB	27 Nov 2019 09:06:23	U	SMPL
K6066-03 - 1	TP-19-S	17194	3.2866	-	7471B	PPB	27 Nov 2019 09:08:39	U	SMPL
K6022-03LX5 - 1		-87	-0.0059	-	7471B	PPB	27 Nov 2019 09:10:55	U	SMPL
K6024-01DLX4 - 1		26055	4.9749	-	7471B	PPB	27 Nov 2019 09:13:09	U	SMPL
K6024-03DLX10 - 1		51645	9.8505	-	7471B	PPB	27 Nov 2019 09:17:45	U	SMPL
CCV46 - 1	CCV46	26784	5.1138	-	7471B	PPB	27 Nov 2019 09:25:59	U	SMPL
CCB46 - 1	CCB46	-24	0.0061	-	7471B	PPB	27 Nov 2019 09:28:15	U	SMPL
K6024-05DLX10 - 1		46932	8.9526	-	7471B	PPB	27 Nov 2019 09:30:30	U	SMPL
K6024-06DLX4 - 1		28862	5.5097	-	7471B	PPB	27 Nov 2019 09:32:46	U	SMPL
L6024-02DLX20 - 1		29655	5.6608	-	7471B	PPB	27 Nov 2019 09:35:04	U	SMPL
K6024-04DLX100 - 1		44490	8.4873	-	7471B	PPB	27 Nov 2019 09:37:21	U	SMPL
CCV47 - 1	CCV47	27636	5.2761	-	7471B	PPB	27 Nov 2019 09:39:38	U	SMPL
CCB47 - 1	CCB47	-17	0.0074	-	7471B	PPB	27 Nov 2019 09:41:57	U	SMPL