


LB106511
INSTRUMENT ID CVI

ISM02.4



Linear

A= 0.0000e+000

B= 2.7223e-004

C= 5.1867e-003

Rho= 0.9990641

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.012	0.012	26	0.000	26					
0.05	0.050										
0.20	0.200	0.196	-0.004	701	0.0 %	701					0
2.50	2.500	2.402	-0.098	8806	0.0 %	8806					-4
5.00	5.000	5.007	0.007	18375	0.0 %	18375					0
7.50	7.500	7.805	0.305	28652	0.0 %	28652					4
10.0	10.000	9.776	-0.224	35893	0.0 %	35893					-2

lb106511
INSTRUMENT ID CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	
0 S0.0	26	-	0	ISM02.4	PPB	11/27/2019 10:43	Std	
0.2 S0.2	701	-	0.2	ISM02.4	PPB	11/27/2019 10:45	Std	
2.5 S2.5	8806	-	2.5	ISM02.4	PPB	11/27/2019 10:47	Std	
5 S5.0	18375	-	5	ISM02.4	PPB	11/27/2019 10:50	Std	
7.5 S7.5	28652	-	7.5	ISM02.4	PPB	11/27/2019 10:52	Std	
10 S10.0	35893	-	10	ISM02.4	PPB	11/27/2019 10:57	Std	
ICV	ICV	14609	3.9823	-	ISM02.4	PPB	11/27/2019 11:00	SMPL
ICB	ICB	-35	-0.0043	-	ISM02.4	PPB	11/27/2019 11:02	SMPL
CCV048	CCV048	18523	5.0478	-	ISM02.4	PPB	11/27/2019 11:04	SMPL
CCB048	CCB048	-61	-0.0114	-	ISM02.4	PPB	11/27/2019 11:06	SMPL
PB125159E	PBW159	11	0.0082	-	ISM02.4	PPB	11/27/2019 11:09	SMPL
K5892-18	MJLLS9	15	0.0093	-	ISM02.4	PPB	11/27/2019 11:11	SMPL
CCV049	CCV049	18620	5.0742	-	ISM02.4	PPB	11/27/2019 11:27	SMPL
CCB049	CCB049	6	0.0068	-	ISM02.4	PPB	11/27/2019 11:29	SMPL

07
11/27/19