



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106513

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/27/2019	12:00
2	CAL2	1	Water	NA	NA	20.1	7.01	11/27/2019	12:02
3	CAL3	1	Water	NA	NA	20.1	10.03	11/27/2019	12:04
4	ICV	1	Water	NA	NA	20.1	7.00	11/27/2019	12:05
5	CCV1	1	Water	NA	NA	20.1	2.02	11/27/2019	12:06
6	K6078-01	1	Water	NA	NA	6.8	2.53	11/27/2019	12:10
7	K6078-01DUP	1	Water	NA	NA	6.9	2.54	11/27/2019	12:11
8	CCV2	1	Water	NA	NA	20.1	12.02	11/27/2019	12:13

✓ 27106513

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K6078 WorkList ID : 133761 Department : Wet-Chemistry Date : 11-27-2019 11:44:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/27/2019	Water	K6078-01	pH	Cool 4 deg C	JACO05	I11	IDW-DECON-WATER-112719	11/27/2019	9040C

Date/Time 11-27-19 11:45
 Received by: Sn
 Relinquished by: CA

Date/Time 11-27-19 12:35
 Received by: sn
 Relinquished by: fr