



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB106523

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	11/27/2019	12:32
2	K6025-02		1	9	100.00	>212.0	>212.0	11/27/2019	12:55
3	K6025-02DUP		1	9	100.00	>212.0	>212.0	11/27/2019	13:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 106523

WorkList Name : fp112719 WorkList ID : 133765 Department : Wet-Chemistry Date : 11-27-2019 11:59:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/09/2019	Water	K6025-02	Flash Point	Cool 4 deg C	SEAC01	I41	GROUNDWATER	11/25/2019	1010A

Date/Time 11/27/19 12:05
 Received by: KS
 Relinquished by: JB

Date/Time 11/27/19 13:25
 Received by: JB
 Relinquished by: KS