



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 12/2/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 11/27/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:40
Out Date: 11/28/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106531

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K6088-01	COBM7	1	1.12	9.56	8.35	85.7	
K6088-02	COBM8	2	1.14	9.77	8.99	91.0	
K6088-03	COBM9	3	1.19	9.78	8.21	81.7	
K6088-04	COBN1	4	1.11	9.86	7.93	77.9	
K6088-05	COBN2	5	1.14	9.66	8.2	82.9	
K6088-06	COBN3	6	1.14	9.75	8.37	84.0	
K6088-07	COBN4	7	1.14	9.88	7.8	76.2	
K6088-08	COBN5	8	1.18	9.71	8.6	87.0	
K6088-09	COBN6	9	1.17	9.98	9.22	91.4	
K6088-10	COBN7	10	1.1	9.85	8.28	82.1	
K6088-11	COBN8	11	1.14	9.69	8.02	80.5	
K6088-12	COBN9	12	1.15	9.98	8.86	87.3	
K6088-13	COBP0	13	1.11	9.76	9.11	92.5	
K6088-14	COBP2	14	1.14	9.78	8.26	82.4	
K6088-15	COBP3	15	1.12	9.64	8.43	85.8	
K6088-16	COBP4	16	1.15	9.94	8.38	82.3	
K6088-17	COBP5	17	1.14	9.77	8.47	84.9	
K6088-18	COBP6	18	1.15	9.72	8.37	84.2	
K6088-19	COBP7	19	1.18	9.92	7.52	72.5	
K6088-20	COBQ8	20	1.15	9.86	8.64	86.0	
K6088-21	COBQ8MS	21	1.15	9.86	8.64	86.0	
K6088-22	COBQ8MSD	22	1.15	9.86	8.64	86.0	
K6088-23	VHBLK01	23	1.00	2.00	2.00	100.0	VHBLK

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$