



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106533

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/27/2019	17:00
2	CAL2	1	Water	NA	NA	20.1	7.01	11/27/2019	17:01
3	CAL3	1	Water	NA	NA	20.1	10.03	11/27/2019	17:02
4	ICV	1	Water	NA	NA	20.2	7.01	11/27/2019	17:05
5	CCV1	1	Water	NA	NA	20.1	2.02	11/27/2019	17:06
6	K6087-01	1	Water	NA	NA	18.6	5.77	11/27/2019	17:08
7	K6087-01DUP	1	Water	NA	NA	18.7	5.78	11/27/2019	17:09
8	K6087-02	1	Water	NA	NA	18.5	4.17	11/27/2019	17:10
9	CCV2	1	Water	NA	NA	20.1	12.01	11/27/2019	17:11

W0106533

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K6087 WATER WorkList ID : 133777 Department : Wet-Chemistry Date : 11-27-2019 15:06:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/27/2019	Water	K6087-01 ✓	pH	Cool 4 deg C	PSEG01	N13	30629	11/27/2019	9040C
11/27/2019	Water	K6087-02 ✓	pH	Cool 4 deg C	PSEG01	N13	30714	11/27/2019	9040C

Date/Time 11-27-19 16:10
 Received by: sf
 Relinquished by: sf

Date/Time 11-27-19 18:00
 Received by: sf
 Relinquished by: sf