



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106535

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/27/2019	17:00
2	CAL2	1	Water	NA	NA	20.1	7.01	11/27/2019	17:01
3	CAL3	1	Water	NA	NA	20.1	10.03	11/27/2019	17:02
4	ICV	1	Water	NA	NA	20.2	7.01	11/27/2019	17:05
5	CCV1	1	Water	NA	NA	20.1	2.02	11/27/2019	17:06
6	K5676-38	1	Solid	20.03	20	21.7	8.90	11/27/2019	17:35
7	K5676-38DUP	1	Solid	20.04	20	21.8	8.91	11/27/2019	17:36
8	K6090-04	1	Solid	20.02	20	21.6	8.42	11/27/2019	17:37
9	CCV2	1	Water	NA	NA	20.1	12.01	11/27/2019	17:39

✓ 106535

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K6090 WorkList ID : 133778 Department : Wet-Chemistry Date : 11-27-2019 15:06:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/27/2019	Solid	K5676-38 -B	Corrosivity	Cool 4 deg C	PSEG05		SU-01-112719	11/27/2019	9045D
11/27/2019	Solid	K6090-04 -B	Corrosivity	Cool 4 deg C	PSEG01	N13	TP-20-S	11/27/2019	9045D

Date/Time 11-27-19 16:00
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 11-27-19 18:00
Received by: [Signature]
Relinquished by: [Signature]