


LB106 544
INSTRUMENT ID CV1

7470A



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.150	-0.150	21	0.000	21				
0.2	0.200	0.224	0.024	1368	0.0 %	1368				
2.5	2.500	2.453	-0.047	9390	0.0 %	9390				
5.0	5.000	5.268	0.268	19523	0.0 %	19523				
7.5	7.500	7.695	0.195	28258	0.0 %	28258				
10.0	10.000	9.709	-0.291	35506	0.0 %	35506				

LB106544 INSTRUMENT ID CV1

Method: 7470A		Operator: Admin		Date of Analysis: 29 Nov 2019 08:08:15				
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50.0	21	-	0.0000	7470A	PPB	29 Nov 2019 08:15:45	S Std
0.2 - 1	50.2	1368	-	0.2000	7470A	PPB	29 Nov 2019 08:17:52	S Std
2.5 - 1	52.5	9390	-	2.5000	7470A	PPB	29 Nov 2019 08:20:00	S Std
5.0 - 1	85.0	19523	-	5.0000	7470A	PPB	29 Nov 2019 08:22:07	S Std
7.5 - 1	87.5	28258	-	7.5000	7470A	PPB	29 Nov 2019 08:24:15	S Std
10.0 - 1	89.0	35506	-	10.0000	7470A	PPB	29 Nov 2019 08:32:33	S Std
ICV12 - 1	ICV12	15801	4.2343	-	7470A	PPB	29 Nov 2019 08:35:06	U SMPL
ICB12 - 1	ICB12	-25	-0.1625	-	7470A	PPB	29 Nov 2019 08:37:12	U SMPL
CCV48 - 1	CCV48	19725	5.3244	-	7470A	PPB	29 Nov 2019 08:40:47	U SMPL
CCB48 - 1	CCB48	-27	-0.1631	-	7470A	PPB	29 Nov 2019 08:42:52	U SMPL
CRA - 1	CRA	1350	0.2195	-	7470A	PPB	29 Nov 2019 08:44:59	U SMPL
HighStd - 1	HighStd	35303	9.6523	-	7470A	PPB	29 Nov 2019 08:47:05	U SMPL
ChkStd - 1	ChkStd	25870	7.0316	-	7470A	PPB	29 Nov 2019 08:49:12	U SMPL
PB125161BL - 1	PBW	-57	-0.1714	-	7470A	PPB	29 Nov 2019 08:51:19	U SMPL
PB125161BS - 1	LCSW	15699	4.2059	-	7470A	PPB	29 Nov 2019 08:53:24	U SMPL
K5986-04 - 1	INK-WASTE	87	-0.1314	-	7470A	PPB	29 Nov 2019 08:55:31	U SMPL
K5986-04DUP - 1	INK-WASTEDUP	68	-0.1367	-	7470A	PPB	29 Nov 2019 08:57:39	U SMPL
K5986-04MS - 1	INK-WASTEMS	21108	5.7087	-	7470A	PPB	29 Nov 2019 08:59:47	U SMPL
K5986-04MSD - 1	INK-WASTEMSD	20564	5.5575	-	7470A	PPB	29 Nov 2019 09:01:55	U SMPL
PB125162BL - 1	PBW	21	-0.1498	-	7470A	PPB	29 Nov 2019 09:04:02	U SMPL
CCV49 - 1	CCV49	19544	5.2742	-	7470A	PPB	29 Nov 2019 09:06:09	U SMPL
CCB49 - 1	CCB49	27	-0.1481	-	7470A	PPB	29 Nov 2019 09:08:16	U SMPL
PB125162BS - 1	LCSW	15421	4.1287	-	7470A	PPB	29 Nov 2019 09:10:22	U SMPL
K5667-04 - 1	NB-07-112519	-9	-0.1581	-	7470A	PPB	29 Nov 2019 09:12:28	U SMPL
K5667-04DUP - 1	NB-07-112519DUP	27	-0.1481	-	7470A	PPB	29 Nov 2019 09:14:34	U SMPL
K5667-04MS - 1	NB-07-112519MS	19361	5.2233	-	7470A	PPB	29 Nov 2019 09:16:41	U SMPL
K5667-04MSD - 1	NB-07-112519MSD	19482	5.2569	-	7470A	PPB	29 Nov 2019 09:18:48	U SMPL
K6010-01 - 1	CALVSOIL1-20191121	-21	-0.1614	-	7470A	PPB	29 Nov 2019 09:20:54	U SMPL
K6013-03 - 1	600436752	94	-0.1295	-	7470A	PPB	29 Nov 2019 09:23:02	U SMPL
K6018-02 - 1	ETGI-359	4	-0.1545	-	7470A	PPB	29 Nov 2019 09:25:10	U SMPL
K6022-04 - 1	TP-15-S	66	-0.1373	-	7470A	PPB	29 Nov 2019 09:27:15	U SMPL
K6022-08 - 1	TP-16-S	45	-0.1431	-	7470A	PPB	29 Nov 2019 09:29:21	U SMPL
CCV50 - 1	CCV50	19583	5.2850	-	7470A	PPB	29 Nov 2019 09:31:28	U SMPL
CCB50 - 1	CCB50	48	-0.1423	-	7470A	PPB	29 Nov 2019 09:33:34	U SMPL
K6022-12 - 1	TP-16-D	45	-0.1431	-	7470A	PPB	29 Nov 2019 09:35:41	U SMPL
K6022-16 - 1	TP-17-S	45	-0.1431	-	7470A	PPB	29 Nov 2019 09:37:48	U SMPL
K6025-01 - 1	SOIL-CUTTINGS	44	-0.1434	-	7470A	PPB	29 Nov 2019 09:39:54	U SMPL
PB125163BL - 1	PBW	37	-0.1453	-	7470A	PPB	29 Nov 2019 09:42:01	U SMPL
PB125163BS - 1	LCSW	15765	4.2243	-	7470A	PPB	29 Nov 2019 09:44:08	U SMPL
K5675-11 - 1	TR-05-112619	44	-0.1434	-	7470A	PPB	29 Nov 2019 09:46:15	U SMPL
K5675-11DUP - 1	TR-05-112619DUP	44	-0.1434	-	7470A	PPB	29 Nov 2019 09:48:23	U SMPL
K5675-11MS - 1	TR-05-112619MS	17484	4.7018	-	7470A	PPB	29 Nov 2019 09:50:30	U SMPL
K5675-11MSD - 1	TR-05-112619MSD	17568	4.7252	-	7470A	PPB	29 Nov 2019 09:52:38	U SMPL
K6054-03 - 1	RBR-50	8	-0.1534	-	7470A	PPB	29 Nov 2019 09:54:45	U SMPL
CCV51 - 1	CCV51	19202	5.1791	-	7470A	PPB	29 Nov 2019 09:56:51	U SMPL
CCB51 - 1	CCB51	6	-0.1539	-	7470A	PPB	29 Nov 2019 09:58:59	U SMPL
K6054-06 - 1	RBR-28	55	-0.1403	-	7470A	PPB	29 Nov 2019 10:01:05	U SMPL
K6066-04 - 1	TP-19-S	97	-0.1286	-	7470A	PPB	29 Nov 2019 10:03:12	U SMPL
PB125172BL - 1	PBW	168	-0.1089	-	7470A	PPB	29 Nov 2019 10:05:18	U SMPL
PB125172BS - 1	LCSW	15996	4.2884	-	7470A	PPB	29 Nov 2019 10:12:09	U SMPL
K5908-10 - 1	B-8D-(12-18)	134	-0.1184	-	7470A	PPB	29 Nov 2019 10:14:15	U SMPL
K5908-10DUP - 1	B-8D-(12-18)DUP	203	-0.0992	-	7470A	PPB	29 Nov 2019 10:16:22	U SMPL
K5908-10MS - 1	B-8D-(12-18)MS	17412	4.6818	-	7470A	PPB	29 Nov 2019 10:18:30	U SMPL
K5908-10MSD - 1	B-8D-(12-18)MSD	16970	4.5590	-	7470A	PPB	29 Nov 2019 10:20:38	U SMPL
K5908-11 - 1	B-8-1-(6-12)	197	-0.1009	-	7470A	PPB	29 Nov 2019 10:22:47	U SMPL
K5908-12 - 1	B-8-2-(6-12)	229	-0.0920	-	7470A	PPB	29 Nov 2019 10:24:53	U SMPL
CCV52 - 1	CCV52	19516	5.2664	-	7470A	PPB	29 Nov 2019 10:27:00	U SMPL
CCB52 - 1	CCB52	8	-0.1534	-	7470A	PPB	29 Nov 2019 10:29:06	U SMPL
K5986-04LX5 - 1		141	-0.1164	-	7470A	PPB	29 Nov 2019 10:31:13	U SMPL
K5667-04LX5 - 1		23	-0.1492	-	7470A	PPB	29 Nov 2019 10:33:20	U SMPL
K5675-11LX5 - 1		39	-0.1448	-	7470A	PPB	29 Nov 2019 10:35:27	U SMPL
K5908-10LX5 - 1		64	-0.1378	-	7470A	PPB	29 Nov 2019 10:37:34	U SMPL
PB125161TB - 1	PB125102TB	40	-0.1445	-	7470A	PPB	29 Nov 2019 10:39:41	U SMPL
PB125162TB - 1	PB125098TB	41	-0.1442	-	7470A	PPB	29 Nov 2019 10:41:48	U SMPL
PB125163TB - 1	PB125139TB	50	-0.1417	-	7470A	PPB	29 Nov 2019 10:43:56	U SMPL
PB125172TB - 1	PB125100TB	37	-0.1453	-	7470A	PPB	29 Nov 2019 10:46:03	U SMPL
CCV53 - 1	CCV53	19498	5.2614	-	7470A	PPB	29 Nov 2019 10:48:12	U SMPL
CCB53 - 1	CCB53	-1	-0.1559	-	7470A	PPB	29 Nov 2019 10:50:19	U SMPL
K5986-04A - 1		15178	4.0612	-	7470A	PPB	29 Nov 2019 11:46:31	U SMPL
K5667-04A - 1		13516	3.5994	-	7470A	PPB	29 Nov 2019 11:48:37	U SMPL
CCV54 - 1	CCV54	17293	4.6488	-	7470A	PPB	29 Nov 2019 11:53:24	U SMPL
CCB54 - 1	CCB54	21	-0.1498	-	7470A	PPB	29 Nov 2019 11:55:30	U SMPL