


LB106553

7471B

INSTRUMENT ID C01



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.032	-0.032	-38	0.000	-38				
0.2	0.200	0.210	0.010	1338	0.0 %	1338				
2.5	2.500	2.512	0.012	14422	0.0 %	14422				
5.0	5.000	5.000	-0.000	28563	0.0 %	28563				
7.5	7.500	7.556	0.056	43094	0.0 %	43094				
10.0	10.000	9.954	-0.046	56725	0.0 %	56725				

LB106553 INSTRUMENT ID CV1

Method: 7471B		Operator: Admin		Date of Analysis: 02 Dec 2019 09:53:37				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50.0	-38	-	0.0000	7471B	PPB	02 Dec 2019 09:56:24	S Std
0.2 - 1	50.2	1338	-	0.2000	7471B	PPB	02 Dec 2019 09:58:41	S Std
2.5 - 1	52.6	14422	-	2.5000	7471B	PPB	02 Dec 2019 10:00:59	S Std
5.0 - 1	55.0	28563	-	5.0000	7471B	PPB	02 Dec 2019 10:03:15	S Std
7.5 - 1	57.5	43094	-	7.5000	7471B	PPB	02 Dec 2019 10:05:32	S Std
10.0 - 1	510.0	56725	-	10.0000	7471B	PPB	02 Dec 2019 10:07:49	S Std
ICV14 - 1	ICV14	22438	3.9222	-	7471B	PPB	02 Dec 2019 10:11:00	U SMPL
ICB14 - 1	ICB14	-17	-0.0283	-	7471B	PPB	02 Dec 2019 10:13:16	U SMPL
CCV59 - 1	CCV59	28082	4.9151	-	7471B	PPB	02 Dec 2019 10:15:32	U SMPL
CCB59 - 1	CCB59	-12	-0.0274	-	7471B	PPB	02 Dec 2019 10:17:48	U SMPL
CRA - 1	CRA	1369	0.2155	-	7471B	PPB	02 Dec 2019 10:20:04	U SMPL
HighStd - 1	HighStd	56509	9.9162	-	7471B	PPB	02 Dec 2019 10:22:19	U SMPL
ChkStd - 1	ChkStd	37182	6.5160	-	7471B	PPB	02 Dec 2019 10:24:34	U SMPL
PB125200BL - 1	PBS	-26	-0.0299	-	7471B	PPB	02 Dec 2019 10:26:52	U SMPL
PB125200BS - 1	LCSS	21263	3.7154	-	7471B	PPB	02 Dec 2019 10:29:09	U SMPL
K5676-35 - 1	SU-01-112719	6146	1.0559	-	7471B	PPB	02 Dec 2019 10:31:26	U SMPL
K5676-35DUP - 1	SU-01-112719DUP	5782	0.9919	-	7471B	PPB	02 Dec 2019 10:33:43	U SMPL
K5676-35MS - 1	SU-01-112719MS	29495	5.1637	-	7471B	PPB	02 Dec 2019 10:38:17	U SMPL
K5676-35MSD - 1	SU-01-112719MSD	29223	5.1158	-	7471B	PPB	02 Dec 2019 10:40:34	U SMPL
K6090-03 - 1	TP-20-S	14204	2.4736	-	7471B	PPB	02 Dec 2019 10:42:49	U SMPL
CCV60 - 1	CCV60	27715	4.8505	-	7471B	PPB	02 Dec 2019 10:45:05	U SMPL
CCB60 - 1	CCB60	-17	-0.0283	-	7471B	PPB	02 Dec 2019 10:47:20	U SMPL
K5676-35LX5 - 1		1139	0.1750	-	7471B	PPB	02 Dec 2019 10:49:37	U SMPL
CCV61 - 1	CCV61	27832	4.8711	-	7471B	PPB	02 Dec 2019 10:51:52	U SMPL
CCB61 - 1	CCB61	28	-0.0204	-	7471B	PPB	02 Dec 2019 10:54:07	U SMPL