



### Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106557

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	12/02/2019	10:35
2	CAL2	1	Water	NA	NA	20.1	7.01	12/02/2019	10:36
3	CAL3	1	Water	NA	NA	20.1	10.03	12/02/2019	10:27
4	ICV	1	Water	NA	NA	20.1	7.00	12/02/2019	10:30
5	CCV1	1	Water	NA	NA	20.2	2.01	12/02/2019	10:31
6	K6093-01	1	Water	NA	NA	16.2	4.70	12/02/2019	10:32
7	K6093-01DUP	1	Water	NA	NA	16.3	4.71	12/02/2019	10:33
8	CCV2	1	Water	NA	NA	20.1	12.02	12/02/2019	10:35

LB 106557

### WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K6093      WorkList ID : 133821      Department : Wet-Chemistry      Date : 12-02-2019 10:26:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/26/2019	Water	K6093-01 - <del>A</del>	pH	Cool 4 deg C	STER08	I61	POOL-WATER	11/26/2019	SM 4500-H B

Date/Time 12-02-19 10:30  
 Received by: Ja  
 Relinquished by: Ja

Date/Time 12-02-19 11:30  
 Received by: Ja  
 Relinquished by: Ja