

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB106560

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP80225
calibration std. phosphate 0.5 ppm	WP80224
calibration std. phosphate 0.3 ppm	WP80223
calibration std. phosphate 0.1 ppm	WP80222
calibration std. phosphate 0.05 ppm	WP80221
calibration std. 0 ppm	WP80220
phosphate ICV std.	WP80227
phosphate CCV std.	WP80226
5N sulfuric acid	WP79628
Combined reagent	WP80229
Phenolphthalein indicator	WP77313
Sodium hydroxide, 1N	WP77314

Intercept: -0.0005Slope: 0.6491Regression: 0.999823

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	12/02/2019	12:06
2	CAL2	0.05	1	50	50	0.033	12/02/2019	12:07
3	CAL3	0.10	1	50	50	0.066	12/02/2019	12:07
4	CAL4	0.30	1	50	50	0.186	12/02/2019	12:08
5	CAL5	0.50	1	50	50	0.330	12/02/2019	12:09
6	CAL6	1.00	1	50	50	0.648	12/02/2019	12:10



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB106560

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.321	0.495	12/02/2019	12:10
2	ICB		1	50	50	0.000	0.001	12/02/2019	12:11
3	CCV1	0.50	1	50	50	0.327	0.505	12/02/2019	12:12
4	CCB1		1	50	50	0.000	0.001	12/02/2019	12:12
5	PB125221BL		1	50	50	0.000	0.001	12/02/2019	12:13
6	PB125221BS	0.50	1	50	50	0.344	0.531	12/02/2019	12:13
7	K6016-01		1	50	50	0.008	0.013	12/02/2019	12:14
8	K6016-03		1	50	50	0.323	0.498	12/02/2019	12:14
9	K6016-04		1	50	50	0.294	0.454	12/02/2019	12:15
10	K6016-07		1	50	50	0.012	0.019	12/02/2019	12:16
11	K6016-08		1	50	50	0.161	0.249	12/02/2019	12:16
12	K6016-09		1	50	50	0.168	0.260	12/02/2019	12:17
13	K6016-09DUP		1	50	50	0.167	0.258	12/02/2019	12:17
14	K6016-09MS	0.50	1	50	50	0.462	0.713	12/02/2019	12:19
15	CCV2	0.50	1	50	50	0.326	0.503	12/02/2019	12:18
16	CCB2		1	50	50	0.000	0.001	12/02/2019	12:19
17	K6016-09MSD	0.50	1	50	50	0.460	0.709	12/02/2019	12:20
18	CCV3	0.50	1	50	50	0.329	0.508	12/02/2019	12:20
19	CCB3		1	50	50	0.000	0.001	12/02/2019	12:21