


LB106582
INSTRUMENT ID CVI

245.1

μ Abs.:

39079



Linear

A= 0.0000e+000

B= 2.5309e-004

C= -2.1000e-002

Rho= 0.9994597

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.016	0.016	146	0.000	146				
0.05	0.050	0.054	0.004	296	0.0 %	296				
0.2	0.200	0.238	0.038	1022	0.0 %	1022				
2.5	2.500	2.428	-0.072	9677	0.0 %	9677				
5.0	5.000	4.885	-0.115	19386	0.0 %	19386				
7.5	7.500	7.759	0.259	30741	0.0 %	30741				
10.0	10.000	9.870	-0.130	39079	0.0 %	39079				

Ib106582 INSTRUMENT ID CV1

Method: 245.1		Operator: Admin		Date of Analysis: 04 Dec 2019 10:29:12					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	146	-	0.0000	245.1	PPB	04 Dec 2019 11:52:53	S	Std
0.05 - 1	50.05	296	-	0.0500	245.1	PPB	04 Dec 2019 11:55:10	S	Std
0.2 - 1	50.2	1022	-	0.2000	245.1	PPB	04 Dec 2019 11:57:27	S	Std
2.5 - 1	52.5	9677	-	2.5000	245.1	PPB	04 Dec 2019 11:59:44	S	Std
5.0 - 1	55.0	19386	-	5.0000	245.1	PPB	04 Dec 2019 12:02:00	S	Std
7.5 - 1	57.5	30741	-	7.5000	245.1	PPB	04 Dec 2019 12:04:16	S	Std
10.0 - 1	510.0	39079	-	10.0000	245.1	PPB	04 Dec 2019 12:06:32	S	Std
ICV16 - 1	ICV16	16283	4.1001	-245.1		PPB	04 Dec 2019 12:54:33	U	SMPL
ICB16 - 1	ICB16	104	0.0053	-245.1		PPB	04 Dec 2019 12:56:48	U	SMPL
CCV66 - 1	CCV66	19273	4.8568	-245.1		PPB	04 Dec 2019 12:59:04	U	SMPL
CCB66 - 1	CCB66	77	-0.0015	-245.1		PPB	04 Dec 2019 13:01:19	U	SMPL
CRA - 1	CRA	289	0.0521	-245.1		PPB	04 Dec 2019 13:03:35	U	SMPL
HighStd - 1	HighStd	38166	9.6385	-245.1		PPB	04 Dec 2019 13:05:50	U	SMPL
ChkStd - 1	ChkStd	27066	6.8292	-245.1		PPB	04 Dec 2019 13:08:06	U	SMPL
PB125233BL - 1	PBW	-24	-0.0271	-245.1		PPB	04 Dec 2019 13:10:22	U	SMPL
PB125233BS - 1	LCSW	15906	4.0047	-245.1		PPB	04 Dec 2019 13:12:38	U	SMPL
K6122-01 - 1	391126065	944	0.2179	-245.1		PPB	04 Dec 2019 13:14:55	U	SMPL
K6122-01DUP - 1	391126065DUP	971	0.2248	-245.1		PPB	04 Dec 2019 13:17:12	U	SMPL
K6122-01MS - 1	391126065MS	16897	4.2555	-245.1		PPB	04 Dec 2019 13:19:58	U	SMPL
K6122-01MSD - 1	391126065MSD	16858	4.2456	-245.1		PPB	04 Dec 2019 13:22:16	U	SMPL
K6123-01 - 1	391127738	952	0.2199	-245.1		PPB	04 Dec 2019 13:24:31	U	SMPL
CCV67 - 1	CCV67	19602	4.9401	-245.1		PPB	04 Dec 2019 13:26:47	U	SMPL
CCB67 - 1	CCB67	-6	-0.0225	-245.1		PPB	04 Dec 2019 13:29:03	U	SMPL
K6124-01 - 1	391127743	1126	0.2640	-245.1		PPB	04 Dec 2019 13:31:18	U	SMPL
K6125-01 - 1	391127003.2	934	0.2154	-245.1		PPB	04 Dec 2019 13:33:34	U	SMPL
K6122.01LX5 - 1		205	0.0309	-245.1		PPB	04 Dec 2019 13:35:49	U	SMPL
K6122-01A - 1		16803	4.2317	-245.1		PPB	04 Dec 2019 13:38:06	U	SMPL
CCV68 - 1	CCV68	19275	4.8573	-245.1		PPB	04 Dec 2019 13:40:22	U	SMPL
CCB68 - 1	CCB68	9	-0.0187	-245.1		PPB	04 Dec 2019 13:42:38	U	SMPL