



## Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106584

Slope : 98.8

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2568         |
| buffer solution pH 7 yellow    | W2447         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2570         |
| buffer solution pH 7 yellow    | W2569         |
| Buffer Solution, PH2 (500ml)   | W2592         |
| Buffer Solution, PH12 (500ml)  | W2616         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.1             | 3.99        | 12/04/2019 | 09:35     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.1             | 7.01        | 12/04/2019 | 09:36     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.1             | 10.03       | 12/04/2019 | 09:37     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.1             | 7.00        | 12/04/2019 | 09:39     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.1             | 2.02        | 12/04/2019 | 09:40     |
| 6   | K6128-01    | 1  | Water  | NA          | NA          | 11.5             | 6.63        | 12/04/2019 | 09:42     |
| 7   | K6128-01DUP | 1  | Water  | NA          | NA          | 11.6             | 6.64        | 12/04/2019 | 09:43     |
| 8   | CCV2        | 1  | Water  | NA          | NA          | 20.2             | 12.02       | 12/04/2019 | 09:45     |

LB106584

## WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K6128

WorkList ID : 133860

Department : Wet-Chemistry

Date : 12-04-2019 09:27:51

| Due Date   | Matrix | Sample   | Test | Preservative | Customer | Storage Location | Customer Sample     | Collect Date | Method      |
|------------|--------|----------|------|--------------|----------|------------------|---------------------|--------------|-------------|
| 12/03/2019 | Water  | K6128-01 | pH   | Cool 4 deg C | APT101   | I63              | WGGM-19-WWW-TSF-042 | 12/03/2019   | SM 4500-H B |

Date/Time 12.04.19 9:30

Received by: Jn

Relinquished by: Jn

Date/Time 12.04.19 10:30

Received by: Jn

Relinquished by: Jn