



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106585

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/03/2019	16:40
2	CAL2	1	Water	NA	NA	20.2	7.01	12/03/2019	16:41
3	CAL3	1	Water	NA	NA	20.2	10.02	12/03/2019	16:42
4	ICV	1	Water	NA	NA	20.2	7.00	12/03/2019	16:43
5	CCV1	1	Water	NA	NA	20.2	2.02	12/03/2019	16:45
6	K6100-02	1	Solid	20.02	20	21.0	6.82	12/03/2019	16:46
7	K6100-02DUP	1	Solid	20.03	20	21.1	6.83	12/03/2019	16:47
8	K6100-04	1	Solid	20.04	20	21.2	5.99	12/03/2019	16:49
9	K6119-01	1	Solid	20.04	20	21.0	8.72	12/03/2019	16:50
10	K6119-03	1	Solid	20.02	20	20.9	7.25	12/03/2019	16:52
11	K6119-05	1	Solid	20.03	20	21.1	7.90	12/03/2019	17:00
12	K6119-07	1	Solid	20.02	20	20.5	8.78	12/03/2019	17:01
13	K6119-09	1	Solid	20.04	20	20.8	7.59	12/03/2019	17:03
14	K6119-11	1	Solid	20.05	20	21.3	7.32	12/03/2019	17:05
15	K6120-02	1	Solid	20.03	20	20.2	8.33	12/03/2019	17:06
16	CCV2	1	Water	NA	NA	20.1	12.02	12/03/2019	17:09
17	K6120-04	1	Solid	20.04	20	20.2	5.17	12/03/2019	17:10
18	K6120-06	1	Solid	20.03	20	20.1	7.37	12/03/2019	17:12
19	K6120-08	1	Solid	20.04	20	20.4	7.92	12/03/2019	17:15
20	CCV3	1	Water	NA	NA	20.1	2.01	12/03/2019	17:16

VB 106585

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K6120 WorkList ID : 133840 Department : Wet-Chemistry Date : 12-03-2019 15:52:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/02/2019	Solid	K6100-02	Corrosivity	Cool 4 deg C	PSEG01	N13	TH-1	12/02/2019	9045D
12/02/2019	Solid	K6100-04	Corrosivity	Cool 4 deg C	PSEG01	N13	TH-9	12/02/2019	9045D
12/02/2019	Solid	K6119-01	Corrosivity	Cool 4 deg C	AKRF02	H32	SB-3-(0-2)	12/02/2019	9045D
12/02/2019	Solid	K6119-03	Corrosivity	Cool 4 deg C	AKRF02	H32	SB-6-(1-3)	12/02/2019	9045D
12/02/2019	Solid	K6119-05	Corrosivity	Cool 4 deg C	AKRF02	H32	SB-4-(0-2)	12/02/2019	9045D
12/02/2019	Solid	K6119-07	Corrosivity	Cool 4 deg C	AKRF02	H32	SB-5-(0-2)	12/02/2019	9045D
12/03/2019	Solid	K6119-09	Corrosivity	Cool 4 deg C	AKRF02	H32	SB-2-(0-2)	12/03/2019	9045D
12/03/2019	Solid	K6119-11	Corrosivity	Cool 4 deg C	AKRF02	H32	SB-1-(0-2)	12/03/2019	9045D
12/03/2019	Solid	K6120-02	Corrosivity	Cool 4 deg C	PSEG01	N13	TH-7	12/03/2019	9045D
12/03/2019	Solid	K6120-04	Corrosivity	Cool 4 deg C	PSEG01	N13	TH-6	12/03/2019	9045D
12/03/2019	Solid	K6120-06	Corrosivity	Cool 4 deg C	PSEG01	N13	TH-4	12/03/2019	9045D
12/03/2019	Solid	K6120-08	Corrosivity	Cool 4 deg C	PSEG01	N13	TH-5	12/03/2019	9045D

Date/Time 12.03.19 16:00
 Received by: Jd
 Relinquished by: [Signature]

Date/Time 12.03.19 18:00
 Received by: [Signature]
 Relinquished by: Jg