



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106589

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/04/2019	13:30
2	CAL2	1	Water	NA	NA	20.2	7.00	12/04/2019	13:31
3	CAL3	1	Water	NA	NA	20.2	10.02	12/04/2019	13:33
4	ICV	1	Water	NA	NA	20.2	7.01	12/04/2019	13:35
5	CCV1	1	Water	NA	NA	20.2	2.02	12/04/2019	13:36
6	K6134-01	1	Water	NA	NA	9.3	6.35	12/04/2019	13:38
7	K6134-01DUP	1	Water	NA	NA	9.4	6.36	12/04/2019	13:39
8	CCV2	1	Water	NA	NA	20.1	12.02	12/04/2019	13:40

28106589

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K6134

WorkList ID : 133873

Department : Wet-Chemistry

Date : 12-04-2019 13:21:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/04/2019	Water	K6134-01-E	pH	Cool 4 deg C	PSEG01	N13	30573	12/04/2019	9040C

Date/Time 12.04.19 13.25
Received by: Jn
Relinquished by: cd

Date/Time 12.04.19 14.30
Received by: cd
Relinquished by: Jn